

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2529

Continuing Calib Date: 07/10/2025

Initial Calibration Date(s): 07/07/2025 07/08/2025

Continuing Calib Time: 08:00

Initial Calibration Time(s): 21:03 04:24

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.48	6.49	6.39	6.59	0.01
Aroclor-1254-2	(2)	6.70	6.70	6.60	6.80	0.00
Aroclor-1254-3	(3)	7.06	7.06	6.96	7.16	0.00
Aroclor-1254-4	(4)	7.35	7.35	7.25	7.45	0.00
Aroclor-1254-5	(5)	7.76	7.76	7.66	7.86	0.00
Tetrachloro-m-xylene		4.49	4.49	4.39	4.59	0.00
Decachlorobiphenyl		10.17	10.17	10.07	10.27	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2529

Continuing Calib Date: 07/10/2025

Initial Calibration Date(s): 07/07/2025

07/08/2025

Continuing Calib Time: 08:00

Initial Calibration Time(s): 21:03

04:24

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1254-1 (1)	5.66	5.66	5.56	5.76	0.00
Aroclor-1254-2 (2)	5.80	5.81	5.71	5.91	0.01
Aroclor-1254-3 (3)	6.21	6.21	6.11	6.31	0.00
Aroclor-1254-4 (4)	6.43	6.44	6.34	6.54	0.01
Aroclor-1254-5 (5)	6.85	6.85	6.75	6.95	0.00
Tetrachloro-m-xylene	3.78	3.78	3.68	3.88	0.00
Decachlorobiphenyl	8.78	8.78	8.68	8.88	0.00