

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/09/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112135.D **Time Analyzed:** 15:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.760	4.661	4.861	547.310	500.000	9.5
Aroclor-1016-2	4.778	4.680	4.880	559.850	500.000	12.0
Aroclor-1016-3	4.835	4.736	4.936	541.660	500.000	8.3
Aroclor-1016-4	4.956	4.856	5.056	551.770	500.000	10.4
Aroclor-1016-5	5.213	5.113	5.313	559.420	500.000	11.9
Aroclor-1260-1	6.250	6.151	6.351	494.590	500.000	-1.1
Aroclor-1260-2	6.439	6.341	6.541	490.920	500.000	-1.8
Aroclor-1260-3	6.805	6.707	6.907	501.480	500.000	0.3
Aroclor-1260-4	7.064	6.967	7.167	509.890	500.000	2.0
Aroclor-1260-5	7.307	7.209	7.409	493.430	500.000	-1.3
Decachlorobiphenyl	8.701	8.603	8.803	49.070	50.000	-1.9
Tetrachloro-m-xylene	3.673	3.573	3.773	58.210	50.000	16.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/09/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112135.D **Time Analyzed:** 15:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.741	4.661	4.861	545.850	500.000	9.2
Aroclor-1016-2	4.760	4.679	4.879	528.330	500.000	5.7
Aroclor-1016-3	4.936	4.854	5.054	533.880	500.000	6.8
Aroclor-1016-4	4.978	4.897	5.097	521.470	500.000	4.3
Aroclor-1016-5	5.190	5.109	5.309	565.670	500.000	13.1
Aroclor-1260-1	6.219	6.138	6.338	488.890	500.000	-2.2
Aroclor-1260-2	6.407	6.326	6.526	494.720	500.000	-1.1
Aroclor-1260-3	6.559	6.478	6.678	504.050	500.000	0.8
Aroclor-1260-4	7.029	6.948	7.148	478.790	500.000	-4.2
Aroclor-1260-5	7.271	7.190	7.390	479.190	500.000	-4.2
Decachlorobiphenyl	8.648	8.568	8.768	48.120	50.000	-3.8
Tetrachloro-m-xylene	3.666	3.584	3.784	56.670	50.000	13.3