

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/09/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112150.D **Time Analyzed:** 21:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.758	4.661	4.861	468.690	500.000	-6.3
Aroclor-1016-2	4.777	4.680	4.880	470.810	500.000	-5.8
Aroclor-1016-3	4.833	4.736	4.936	465.050	500.000	-7.0
Aroclor-1016-4	4.953	4.856	5.056	477.440	500.000	-4.5
Aroclor-1016-5	5.209	5.113	5.313	456.730	500.000	-8.7
Aroclor-1260-1	6.248	6.151	6.351	481.940	500.000	-3.6
Aroclor-1260-2	6.437	6.341	6.541	479.060	500.000	-4.2
Aroclor-1260-3	6.804	6.707	6.907	443.870	500.000	-11.2
Aroclor-1260-4	7.063	6.967	7.167	485.650	500.000	-2.9
Aroclor-1260-5	7.306	7.209	7.409	480.500	500.000	-3.9
Decachlorobiphenyl	8.699	8.603	8.803	49.530	50.000	-0.9
Tetrachloro-m-xylene	3.671	3.573	3.773	47.450	50.000	-5.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/09/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112150.D **Time Analyzed:** 21:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.741	4.661	4.861	443.040	500.000	-11.4
Aroclor-1016-2	4.759	4.679	4.879	423.560	500.000	-15.3
Aroclor-1016-3	4.935	4.854	5.054	450.350	500.000	-9.9
Aroclor-1016-4	4.977	4.897	5.097	437.130	500.000	-12.6
Aroclor-1016-5	5.189	5.109	5.309	499.940	500.000	0.0
Aroclor-1260-1	6.218	6.138	6.338	453.540	500.000	-9.3
Aroclor-1260-2	6.407	6.326	6.526	463.840	500.000	-7.2
Aroclor-1260-3	6.558	6.478	6.678	467.640	500.000	-6.5
Aroclor-1260-4	7.028	6.948	7.148	454.270	500.000	-9.1
Aroclor-1260-5	7.270	7.190	7.390	463.030	500.000	-7.4
Decachlorobiphenyl	8.647	8.568	8.768	47.890	50.000	-4.2
Tetrachloro-m-xylene	3.666	3.584	3.784	45.880	50.000	-8.2