

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112165.D **Time Analyzed:** 03:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.760	4.661	4.861	481.070	500.000	-3.8
Aroclor-1016-2	4.779	4.680	4.880	492.650	500.000	-1.5
Aroclor-1016-3	4.836	4.736	4.936	477.450	500.000	-4.5
Aroclor-1016-4	4.955	4.856	5.056	483.930	500.000	-3.2
Aroclor-1016-5	5.212	5.113	5.313	451.060	500.000	-9.8
Aroclor-1260-1	6.250	6.151	6.351	437.900	500.000	-12.4
Aroclor-1260-2	6.440	6.341	6.541	447.270	500.000	-10.5
Aroclor-1260-3	6.806	6.707	6.907	442.910	500.000	-11.4
Aroclor-1260-4	7.067	6.967	7.167	465.480	500.000	-6.9
Aroclor-1260-5	7.309	7.209	7.409	447.400	500.000	-10.5
Decachlorobiphenyl	8.703	8.603	8.803	46.070	50.000	-7.9
Tetrachloro-m-xylene	3.673	3.573	3.773	50.290	50.000	0.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112165.D **Time Analyzed:** 03:11

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.741	4.661	4.861	479.680	500.000	-4.1
Aroclor-1016-2	4.761	4.679	4.879	429.670	500.000	-14.1
Aroclor-1016-3	4.935	4.854	5.054	457.340	500.000	-8.5
Aroclor-1016-4	4.978	4.897	5.097	436.680	500.000	-12.7
Aroclor-1016-5	5.191	5.109	5.309	523.480	500.000	4.7
Aroclor-1260-1	6.220	6.138	6.338	450.870	500.000	-9.8
Aroclor-1260-2	6.408	6.326	6.526	451.460	500.000	-9.7
Aroclor-1260-3	6.559	6.478	6.678	455.420	500.000	-8.9
Aroclor-1260-4	7.029	6.948	7.148	443.960	500.000	-11.2
Aroclor-1260-5	7.271	7.190	7.390	446.370	500.000	-10.7
Decachlorobiphenyl	8.649	8.568	8.768	46.990	50.000	-6.0
Tetrachloro-m-xylene	3.666	3.584	3.784	47.770	50.000	-4.5