

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/09/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073640.D **Time Analyzed:** 15:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.639	5.539	5.739	424.810	500.000	-15.0
Aroclor-1016-2	5.660	5.560	5.760	436.050	500.000	-12.8
Aroclor-1016-3	5.722	5.622	5.822	438.140	500.000	-12.4
Aroclor-1016-4	5.820	5.720	5.920	445.700	500.000	-10.9
Aroclor-1016-5	6.112	6.012	6.212	452.970	500.000	-9.4
Aroclor-1260-1	7.229	7.129	7.329	416.990	500.000	-16.6
Aroclor-1260-2	7.481	7.383	7.583	367.360	500.000	-26.5
Aroclor-1260-3	7.840	7.740	7.940	402.830	500.000	-19.4
Aroclor-1260-4	8.065	7.964	8.164	408.010	500.000	-18.4
Aroclor-1260-5	8.382	8.283	8.483	414.130	500.000	-17.2
Decachlorobiphenyl	10.179	10.076	10.276	42.740	50.000	-14.5
Tetrachloro-m-xylene	4.488	4.388	4.588	45.900	50.000	-8.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/09/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073640.D **Time Analyzed:** 15:20

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.858	4.758	4.958	457.520	500.000	-8.5
Aroclor-1016-2	4.876	4.775	4.975	453.260	500.000	-9.3
Aroclor-1016-3	5.053	4.951	5.151	458.820	500.000	-8.2
Aroclor-1016-4	5.094	4.993	5.193	454.810	500.000	-9.0
Aroclor-1016-5	5.308	5.207	5.407	472.620	500.000	-5.5
Aroclor-1260-1	6.338	6.237	6.437	443.640	500.000	-11.3
Aroclor-1260-2	6.526	6.425	6.625	435.900	500.000	-12.8
Aroclor-1260-3	6.678	6.577	6.777	428.880	500.000	-14.2
Aroclor-1260-4	7.148	7.046	7.246	447.130	500.000	-10.6
Aroclor-1260-5	7.391	7.289	7.489	432.670	500.000	-13.5
Decachlorobiphenyl	8.782	8.680	8.880	47.820	50.000	-4.4
Tetrachloro-m-xylene	3.781	3.678	3.878	46.430	50.000	-7.1