

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/09/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073655.D **Time Analyzed:** 20:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.639	5.539	5.739	486.000	500.000	-2.8
Aroclor-1016-2	5.660	5.560	5.760	498.610	500.000	-0.3
Aroclor-1016-3	5.723	5.622	5.822	497.930	500.000	-0.4
Aroclor-1016-4	5.820	5.720	5.920	509.260	500.000	1.9
Aroclor-1016-5	6.112	6.012	6.212	526.740	500.000	5.3
Aroclor-1260-1	7.229	7.129	7.329	501.500	500.000	0.3
Aroclor-1260-2	7.483	7.383	7.583	435.920	500.000	-12.8
Aroclor-1260-3	7.840	7.740	7.940	493.130	500.000	-1.4
Aroclor-1260-4	8.064	7.964	8.164	514.920	500.000	3.0
Aroclor-1260-5	8.382	8.283	8.483	507.030	500.000	1.4
Decachlorobiphenyl	10.177	10.076	10.276	55.510	50.000	11.0
Tetrachloro-m-xylene	4.488	4.388	4.588	51.800	50.000	3.6

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL05 **Date Analyzed:** 07/09/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073655.D **Time Analyzed:** 20:17

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.857	4.758	4.958	510.430	500.000	2.1
Aroclor-1016-2	4.875	4.775	4.975	511.730	500.000	2.3
Aroclor-1016-3	5.052	4.951	5.151	520.490	500.000	4.1
Aroclor-1016-4	5.093	4.993	5.193	519.190	500.000	3.8
Aroclor-1016-5	5.307	5.207	5.407	532.590	500.000	6.5
Aroclor-1260-1	6.337	6.237	6.437	501.490	500.000	0.3
Aroclor-1260-2	6.525	6.425	6.625	496.300	500.000	-0.7
Aroclor-1260-3	6.677	6.577	6.777	495.780	500.000	-0.8
Aroclor-1260-4	7.147	7.046	7.246	519.020	500.000	3.8
Aroclor-1260-5	7.389	7.289	7.489	521.930	500.000	4.4
Decachlorobiphenyl	8.780	8.680	8.880	56.570	50.000	13.1
Tetrachloro-m-xylene	3.779	3.678	3.878	50.850	50.000	1.7