

Surrogate Summary

SDG No.: **Q2529**

Client: **CDM Smith**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PO112087.D	PIBLK-PO112087.D	Tetrachloro-m-xyl	1	20	19.4	97		60	140
		Decachlorobiphen	1	20	18.8	94		60	140
		Tetrachloro-m-xyl	2	20	20.0	100		60	140
		Decachlorobiphen	2	20	19.3	97		60	140
I.BLK-PO112139.D	PIBLK-PO112139.D	Tetrachloro-m-xyl	1	20	17.9	89		60	140
		Decachlorobiphen	1	20	19.4	97		60	140
		Tetrachloro-m-xyl	2	20	17.6	88		60	140
		Decachlorobiphen	2	20	19.3	96		60	140
Q2529-01	TP-91	Tetrachloro-m-xyl	1	20	17.9	89		32	144
		Decachlorobiphen	1	20	13.7	68		32	175
		Tetrachloro-m-xyl	2	20	17.3	87		32	144
		Decachlorobiphen	2	20	13.5	68		32	175
Q2529-02	TP-80	Tetrachloro-m-xyl	1	20	18.7	94		32	144
		Decachlorobiphen	1	20	15.5	78		32	175
		Tetrachloro-m-xyl	2	20	18.3	91		32	144
		Decachlorobiphen	2	20	15.6	78		32	175
Q2529-03	TP-79	Tetrachloro-m-xyl	1	20	19.0	95		32	144
		Decachlorobiphen	1	20	16.1	81		32	175
		Tetrachloro-m-xyl	2	20	18.7	93		32	144
		Decachlorobiphen	2	20	16.1	81		32	175
Q2529-04	TP-95	Tetrachloro-m-xyl	1	20	18.7	93		32	144
		Decachlorobiphen	1	20	12.8	64		32	175
		Tetrachloro-m-xyl	2	20	18.4	92		32	144
		Decachlorobiphen	2	20	12.8	64		32	175
Q2529-05	TP-98	Tetrachloro-m-xyl	1	20	21.5	108		32	144
		Decachlorobiphen	1	20	17.1	86		32	175
		Tetrachloro-m-xyl	2	20	21.0	105		32	144
		Decachlorobiphen	2	20	16.9	84		32	175
Q2529-06	TP-102	Tetrachloro-m-xyl	1	20	19.4	97		32	144
		Decachlorobiphen	1	20	13.8	69		32	175
		Tetrachloro-m-xyl	2	20	18.8	94		32	144
		Decachlorobiphen	2	20	16.0	80		32	175
Q2529-07	TP-101	Tetrachloro-m-xyl	1	20	19.7	99		32	144
		Decachlorobiphen	1	20	13.8	69		32	175
		Tetrachloro-m-xyl	2	20	18.9	95		32	144
		Decachlorobiphen	2	20	15.4	77		32	175
I.BLK-PO112154.D	PIBLK-PO112154.D	Tetrachloro-m-xyl	1	20	20.1	100		60	140
		Decachlorobiphen	1	20	20.6	103		60	140
		Tetrachloro-m-xyl	2	20	19.6	98		60	140
		Decachlorobiphen	2	20	20.1	101		60	140
Q2529-08	TP-89	Tetrachloro-m-xyl	1	20	17.0	85		32	144

Surrogate Summary

SDG No.: Q2529

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
Q2529-08	TP-89	Decachlorobiphen	1	20	16.2	81		32	175
		Tetrachloro-m-xyl	2	20	16.6	83		32	144
Q2529-09	TP-33	Decachlorobiphen	2	20	16.1	81		32	175
		Tetrachloro-m-xyl	1	20	18.5	92		32	144
		Decachlorobiphen	1	20	19.2	96		32	175
		Tetrachloro-m-xyl	2	20	17.9	89		32	144
		Decachlorobiphen	2	20	19.0	95		32	175
Q2529-10	TP-30	Tetrachloro-m-xyl	1	20	18.7	94		32	144
		Decachlorobiphen	1	20	19.0	95		32	175
		Tetrachloro-m-xyl	2	20	18.2	91		32	144
		Decachlorobiphen	2	20	18.6	93		32	175
I.BLK-PO112169.D	PIBLK-PO112169.D	Tetrachloro-m-xyl	1	20	22.1	110		60	140
		Decachlorobiphen	1	20	20.3	102		60	140
		Tetrachloro-m-xyl	2	20	21.3	106		60	140
		Decachlorobiphen	2	20	20.2	101		60	140
I.BLK-PP073553.D	PIBLK-PP073553.D	Tetrachloro-m-xyl	1	20	16.2	81		60	140
		Decachlorobiphen	1	20	17.3	87		60	140
		Tetrachloro-m-xyl	2	20	16.2	81		60	140
		Decachlorobiphen	2	20	17.1	85		60	140
I.BLK-PP073644.D	PIBLK-PP073644.D	Tetrachloro-m-xyl	1	20	18.3	92		60	140
		Decachlorobiphen	1	20	17.7	89		60	140
		Tetrachloro-m-xyl	2	20	18.5	93		60	140
		Decachlorobiphen	2	20	19.7	99		60	140
PB168764BL	PB168764BL	Tetrachloro-m-xyl	1	20	18.1	91		32	144
		Decachlorobiphen	1	20	18.1	90		32	175
		Tetrachloro-m-xyl	2	20	18.5	92		32	144
		Decachlorobiphen	2	20	19.8	99		32	175
PB168764BS	PB168764BS	Tetrachloro-m-xyl	1	20	18.6	93		32	144
		Decachlorobiphen	1	20	18.2	91		32	175
		Tetrachloro-m-xyl	2	20	17.4	87		32	144
		Decachlorobiphen	2	20	20.7	104		32	175
Q2517-01MS	TP-14MS	Tetrachloro-m-xyl	1	20	21.5	107		32	144
		Decachlorobiphen	1	20	20.6	103		32	175
		Tetrachloro-m-xyl	2	20	20.5	102		32	144
		Decachlorobiphen	2	20	22.6	113		32	175
Q2517-01MSD	TP-14MSD	Tetrachloro-m-xyl	1	20	21.3	106		32	144
		Decachlorobiphen	1	20	19.9	100		32	175
		Tetrachloro-m-xyl	2	20	20.7	104		32	144
		Decachlorobiphen	2	20	22.9	115		32	175
I.BLK-PP073659.D	PIBLK-PP073659.D	Tetrachloro-m-xyl	1	20	19.1	96		60	140
		Decachlorobiphen	1	20	19.0	95		60	140

Surrogate Summary

SDG No.: Q2529

Client: CDM Smith

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP073659.D	PIBLK-PP073659.D	Tetrachloro-m-xyl	2	20	18.7	93		60	140
		Decachlorobiphen	2	20	20.9	104		60	140