

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089401.D **Time Analyzed:** 13:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.709	6.605	6.805	47.640	50.000	-4.7
4,4'-DDE	6.200	6.096	6.296	47.460	50.000	-5.1
4,4'-DDT	7.025	6.921	7.121	47.840	50.000	-4.3
Aldrin	5.275	5.171	5.371	47.380	50.000	-5.2
alpha-BHC	4.003	3.899	4.099	47.780	50.000	-4.4
alpha-Chlordane	6.030	5.927	6.127	47.010	50.000	-6.0
beta-BHC	4.520	4.415	4.615	46.410	50.000	-7.2
Decachlorobiphenyl	9.076	8.972	9.172	52.120	50.000	4.2
delta-BHC	4.768	4.664	4.864	49.460	50.000	-1.1
Dieldrin	6.350	6.247	6.447	47.540	50.000	-4.9
Endosulfan I	6.078	5.975	6.175	47.010	50.000	-6.0
Endosulfan II	6.790	6.686	6.886	47.510	50.000	-5.0
Endosulfan sulfate	7.153	7.049	7.249	46.870	50.000	-6.3
Endrin	6.578	6.475	6.675	46.110	50.000	-7.8
Endrin aldehyde	6.919	6.815	7.015	46.450	50.000	-7.1
Endrin ketone	7.634	7.530	7.730	48.060	50.000	-3.9
gamma-BHC (Lindane)	4.334	4.230	4.430	47.510	50.000	-5.0
gamma-Chlordane	5.950	5.846	6.046	47.070	50.000	-5.9
Heptachlor	4.934	4.829	5.029	48.280	50.000	-3.4
Heptachlor epoxide	5.695	5.591	5.791	46.180	50.000	-7.6
Methoxychlor	7.497	7.393	7.593	46.680	50.000	-6.6
Tetrachloro-m-xylene	3.554	3.450	3.650	46.300	50.000	-7.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089401.D **Time Analyzed:** 13:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.928	5.830	6.030	48.630	50.000	-2.7
4,4'-DDE	5.373	5.275	5.475	48.820	50.000	-2.4
4,4'-DDT	6.182	6.084	6.284	48.690	50.000	-2.6
Aldrin	4.367	4.269	4.469	48.800	50.000	-2.4
alpha-BHC	3.391	3.293	3.493	48.860	50.000	-2.3
alpha-Chlordane	5.188	5.091	5.291	48.670	50.000	-2.7
beta-BHC	4.023	3.926	4.126	47.840	50.000	-4.3
Decachlorobiphenyl	8.070	7.972	8.172	55.930	50.000	11.9
delta-BHC	4.260	4.162	4.362	48.670	50.000	-2.7
Dieldrin	5.511	5.413	5.613	48.480	50.000	-3.0
Endosulfan I	5.245	5.147	5.347	49.330	50.000	-1.3
Endosulfan II	6.079	5.981	6.181	48.430	50.000	-3.1
Endosulfan sulfate	6.481	6.383	6.583	48.620	50.000	-2.8
Endrin	5.787	5.689	5.889	46.580	50.000	-6.8
Endrin aldehyde	6.257	6.159	6.359	47.770	50.000	-4.5
Endrin ketone	6.990	6.892	7.092	49.090	50.000	-1.8
gamma-BHC (Lindane)	3.726	3.630	3.830	48.780	50.000	-2.4
gamma-Chlordane	5.124	5.026	5.226	48.660	50.000	-2.7
Heptachlor	4.081	3.983	4.183	47.720	50.000	-4.6
Heptachlor epoxide	4.871	4.773	4.973	49.840	50.000	-0.3
Methoxychlor	6.752	6.655	6.855	46.240	50.000	-7.5
Tetrachloro-m-xylene	2.878	2.780	2.980	48.620	50.000	-2.8