

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089413.D **Time Analyzed:** 17:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.708	6.605	6.805	45.880	50.000	-8.2
4,4'-DDE	6.199	6.096	6.296	45.280	50.000	-9.4
4,4'-DDT	7.024	6.921	7.121	43.960	50.000	-12.1
Aldrin	5.275	5.171	5.371	46.570	50.000	-6.9
alpha-BHC	4.003	3.899	4.099	47.370	50.000	-5.3
alpha-Chlordane	6.030	5.927	6.127	45.680	50.000	-8.6
beta-BHC	4.519	4.415	4.615	46.220	50.000	-7.6
Decachlorobiphenyl	9.076	8.972	9.172	50.570	50.000	1.1
delta-BHC	4.768	4.664	4.864	49.070	50.000	-1.9
Dieldrin	6.351	6.247	6.447	45.940	50.000	-8.1
Endosulfan I	6.078	5.975	6.175	45.610	50.000	-8.8
Endosulfan II	6.790	6.686	6.886	45.590	50.000	-8.8
Endosulfan sulfate	7.152	7.049	7.249	44.200	50.000	-11.6
Endrin	6.578	6.475	6.675	44.880	50.000	-10.2
Endrin aldehyde	6.918	6.815	7.015	43.500	50.000	-13.0
Endrin ketone	7.634	7.530	7.730	45.050	50.000	-9.9
gamma-BHC (Lindane)	4.334	4.230	4.430	46.980	50.000	-6.0
gamma-Chlordane	5.949	5.846	6.046	45.780	50.000	-8.4
Heptachlor	4.933	4.829	5.029	47.130	50.000	-5.7
Heptachlor epoxide	5.694	5.591	5.791	45.010	50.000	-10.0
Methoxychlor	7.496	7.393	7.593	42.190	50.000	-15.6
Tetrachloro-m-xylene	3.554	3.450	3.650	45.900	50.000	-8.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089413.D **Time Analyzed:** 17:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.929	5.830	6.030	47.540	50.000	-4.9
4,4'-DDE	5.375	5.275	5.475	47.860	50.000	-4.3
4,4'-DDT	6.183	6.084	6.284	46.570	50.000	-6.9
Aldrin	4.368	4.269	4.469	48.690	50.000	-2.6
alpha-BHC	3.393	3.293	3.493	49.090	50.000	-1.8
alpha-Chlordane	5.190	5.091	5.291	48.060	50.000	-3.9
beta-BHC	4.025	3.926	4.126	48.020	50.000	-4.0
Decachlorobiphenyl	8.073	7.972	8.172	52.410	50.000	4.8
delta-BHC	4.262	4.162	4.362	48.870	50.000	-2.3
Dieldrin	5.511	5.413	5.613	47.660	50.000	-4.7
Endosulfan I	5.246	5.147	5.347	48.310	50.000	-3.4
Endosulfan II	6.080	5.981	6.181	46.980	50.000	-6.0
Endosulfan sulfate	6.482	6.383	6.583	46.970	50.000	-6.1
Endrin	5.789	5.689	5.889	46.740	50.000	-6.5
Endrin aldehyde	6.258	6.159	6.359	46.010	50.000	-8.0
Endrin ketone	6.992	6.892	7.092	46.480	50.000	-7.0
gamma-BHC (Lindane)	3.728	3.630	3.830	48.710	50.000	-2.6
gamma-Chlordane	5.125	5.026	5.226	48.280	50.000	-3.4
Heptachlor	4.082	3.983	4.183	47.870	50.000	-4.3
Heptachlor epoxide	4.871	4.773	4.973	48.190	50.000	-3.6
Methoxychlor	6.754	6.655	6.855	44.730	50.000	-10.5
Tetrachloro-m-xylene	2.880	2.780	2.980	48.850	50.000	-2.3