

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089425.D **Time Analyzed:** 20:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	46.270	50.000	-7.5
4,4'-DDE	6.195	6.096	6.296	45.890	50.000	-8.2
4,4'-DDT	7.020	6.921	7.121	44.560	50.000	-10.9
Aldrin	5.270	5.171	5.371	47.300	50.000	-5.4
alpha-BHC	3.999	3.899	4.099	48.090	50.000	-3.8
alpha-Chlordane	6.026	5.927	6.127	46.070	50.000	-7.9
beta-BHC	4.515	4.415	4.615	46.700	50.000	-6.6
Decachlorobiphenyl	9.071	8.972	9.172	51.740	50.000	3.5
delta-BHC	4.764	4.664	4.864	50.340	50.000	0.7
Dieldrin	6.345	6.247	6.447	46.270	50.000	-7.5
Endosulfan I	6.074	5.975	6.175	46.060	50.000	-7.9
Endosulfan II	6.785	6.686	6.886	45.820	50.000	-8.4
Endosulfan sulfate	7.148	7.049	7.249	44.990	50.000	-10.0
Endrin	6.573	6.475	6.675	44.280	50.000	-11.4
Endrin aldehyde	6.914	6.815	7.015	44.150	50.000	-11.7
Endrin ketone	7.629	7.530	7.730	45.340	50.000	-9.3
gamma-BHC (Lindane)	4.330	4.230	4.430	47.870	50.000	-4.3
gamma-Chlordane	5.945	5.846	6.046	46.360	50.000	-7.3
Heptachlor	4.929	4.829	5.029	47.400	50.000	-5.2
Heptachlor epoxide	5.690	5.591	5.791	46.080	50.000	-7.8
Methoxychlor	7.491	7.393	7.593	42.620	50.000	-14.8
Tetrachloro-m-xylene	3.550	3.450	3.650	46.860	50.000	-6.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2529
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/09/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089425.D **Time Analyzed:** 20:09

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.928	5.830	6.030	49.490	50.000	-1.0
4,4'-DDE	5.374	5.275	5.475	49.540	50.000	-0.9
4,4'-DDT	6.182	6.084	6.284	47.780	50.000	-4.4
Aldrin	4.368	4.269	4.469	50.210	50.000	0.4
alpha-BHC	3.393	3.293	3.493	50.410	50.000	0.8
alpha-Chlordane	5.189	5.091	5.291	49.500	50.000	-1.0
beta-BHC	4.025	3.926	4.126	49.730	50.000	-0.5
Decachlorobiphenyl	8.071	7.972	8.172	54.420	50.000	8.8
delta-BHC	4.261	4.162	4.362	50.380	50.000	0.8
Dieldrin	5.512	5.413	5.613	49.390	50.000	-1.2
Endosulfan I	5.246	5.147	5.347	49.970	50.000	-0.1
Endosulfan II	6.079	5.981	6.181	48.500	50.000	-3.0
Endosulfan sulfate	6.481	6.383	6.583	48.470	50.000	-3.1
Endrin	5.788	5.689	5.889	47.750	50.000	-4.5
Endrin aldehyde	6.258	6.159	6.359	47.430	50.000	-5.1
Endrin ketone	6.990	6.892	7.092	48.230	50.000	-3.5
gamma-BHC (Lindane)	3.728	3.630	3.830	50.320	50.000	0.6
gamma-Chlordane	5.124	5.026	5.226	49.780	50.000	-0.4
Heptachlor	4.082	3.983	4.183	49.380	50.000	-1.2
Heptachlor epoxide	4.872	4.773	4.973	51.060	50.000	2.1
Methoxychlor	6.753	6.655	6.855	45.850	50.000	-8.3
Tetrachloro-m-xylene	2.881	2.780	2.980	50.370	50.000	0.7