

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070925\
 Data File : BM050378.D
 Acq On : 08 Jul 2025 13:19
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005

Quant Time: Jul 08 18:19:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM070925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 08 17:58:12 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.869	152	332042	20.000	ng	0.00
21) Naphthalene-d8	10.663	136	1221582	20.000	ng	0.00
39) Acenaphthene-d10	14.492	164	768894	20.000	ng	0.00
64) Phenanthrene-d10	17.227	188	1453955	20.000	ng	0.00
76) Chrysene-d12	21.450	240	1431554	20.000	ng	0.00
86) Perylene-d12	24.491	264	1410719	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	186878	9.688	ng	0.00
7) Phenol-d6	7.016	99	226950	9.333	ng	0.00
23) Nitrobenzene-d5	9.016	82	221107	9.234	ng	0.00
42) 2,4,6-Tribromophenol	15.968	330	69925	7.485	ng	0.00
45) 2-Fluorobiphenyl	13.116	172	604282	9.705	ng	0.00
79) Terphenyl-d14	19.839	244	783562	9.630	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.334	88	41862	5.278	ng	99
3) Pyridine	3.734	79	103718	4.981	ng	94
6) Aniline	7.187	93	148080	4.745	ng	99
8) 2-Chlorophenol	7.428	128	94226	4.632	ng	98
10) Phenol	7.045	94	121196	4.778	ng	100
11) bis(2-Chloroethyl)ether	7.281	93	99772	4.912	ng	97
12) 1,3-Dichlorobenzene	7.757	146	124875	5.049	ng	99
13) 1,4-Dichlorobenzene	7.898	146	128887	5.103	ng	98
14) 1,2-Dichlorobenzene	8.216	146	122972	5.086	ng	99
16) 2,2'-oxybis(1-Chloropr...	8.392	45	152357	5.092	ng	99
17) 2-Methylphenol	8.298	107	76571	4.655	ng	97
18) Hexachloroethane	8.951	117	42990	4.840	ng	97
19) n-Nitroso-di-n-propyla...	8.663	70	66842	4.537	ng	96
22) Acetophenone	8.681	105	150473	4.927	ng	# 98
24) Nitrobenzene	9.057	77	101661	4.758	ng	97
25) Isophorone	9.586	82	174388	4.488	ng	99
26) 2-Nitrophenol	9.769	139	32805	3.767	ng	91
27) 2,4-Dimethylphenol	9.828	122	90013	4.858	ng	99
28) bis(2-Chloroethoxy)met...	10.069	93	124150	4.822	ng	97
29) 2,4-Dichlorophenol	10.304	162	84485	4.438	ng	97
30) 1,2,4-Trichlorobenzene	10.522	180	112916	4.958	ng	98
31) Naphthalene	10.710	128	315550	5.086	ng	99
33) 4-Chloroaniline	10.810	127	121575	4.635	ng	98
34) Hexachlorobutadiene	11.010	225	67689	4.870	ng	98
36) 4-Chloro-3-methylphenol	11.927	107	78716	4.428	ng	99
37) 2-Methylnaphthalene	12.322	142	186482	4.760	ng	98
38) 1-Methylnaphthalene	12.539	142	198866	4.796	ng	99
40) 1,2,4,5-Tetrachloroben...	12.686	216	113496	4.642	ng	99
43) 2,4,6-Trichlorophenol	12.922	196	63524	4.221	ng	100
44) 2,4,5-Trichlorophenol	12.992	196	68958	4.195	ng	98
46) 1,1'-Biphenyl	13.327	154	281496	4.934	ng	99
47) 2-Chloronaphthalene	13.369	162	223672	4.918	ng	99
48) 2-Nitroaniline	13.563	65	37953	3.738	ng	97
49) Acenaphthylene	14.216	152	316292	4.601	ng	98
50) Dimethylphthalate	13.945	163	254362	4.805	ng	99
51) 2,6-Dinitrotoluene	14.057	165	37711	3.710	ng	88

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070925\
 Data File : BM050378.D
 Acq On : 08 Jul 2025 13:19
 Operator : RC/JU
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005

Quant Time: Jul 08 18:19:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM070925.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 08 17:58:12 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
52) Acenaphthene	14.557	154	208840	4.779	ng	97
53) 3-Nitroaniline	14.380	138	39220	3.628	ng #	85
55) Dibenzofuran	14.886	168	335867	4.988	ng	99
57) 2,4-Dinitrotoluene	14.839	165	44983	5.513	ng #	87
58) Fluorene	15.533	166	259083	4.671	ng	98
59) 2,3,4,6-Tetrachlorophenol	15.110	232	57532	4.158	ng	95
60) Diethylphthalate	15.310	149	233863	4.625	ng	98
61) 4-Chlorophenyl-phenyle...	15.533	204	132153	4.557	ng	94
62) 4-Nitroaniline	15.533	138	36922	5.398	ng	80
63) Azobenzene	15.821	77	208493	4.490	ng	96
66) n-Nitrosodiphenylamine	15.739	169	207747	4.566	ng	98
67) 4-Bromophenyl-phenylether	16.421	248	71279	4.450	ng	92
68) Hexachlorobenzene	16.539	284	86781	4.587	ng	99
69) Atrazine	16.686	200	60444	4.046	ng	97
71) Phenanthrene	17.268	178	406647	4.943	ng	99
72) Anthracene	17.357	178	373674	4.563	ng	99
73) Carbazole	17.621	167	341196	4.605	ng	99
74) Di-n-butylphthalate	18.198	149	346654	4.274	ng	99
75) Fluoranthene	19.274	202	395390	4.421	ng	98
78) Pyrene	19.633	202	428758	4.676	ng	100
80) Butylbenzylphthalate	20.533	149	113580	3.757	ng	94
81) Benzo(a)anthracene	21.433	228	430980	4.620	ng	100
83) Chrysene	21.497	228	415885	4.727	ng	98
84) Bis(2-ethylhexyl)phtha...	21.368	149	178701	3.724	ng	98
87) Indeno(1,2,3-cd)pyrene	27.921	276	426152	4.249	ng	95
88) Benzo(b)fluoranthene	23.527	252	384470	4.431	ng	98
89) Benzo(k)fluoranthene	23.591	252	389878	4.330	ng #	97
90) Benzo(a)pyrene	24.344	252	339308	4.177	ng	97
91) Dibenzo(a,h)anthracene	27.991	278	360576	4.269	ng	98
92) Benzo(g,h,i)perylene	28.985	276	349804	4.382	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070925\
Data File : BM050378.D
Acq On : 08 Jul 2025 13:19
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Quant Time: Jul 08 18:19:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM070925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 08 17:58:12 2025
Response via : Initial Calibration




Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM070925\
Data File : BM050378.D
Acq On : 08 Jul 2025 13:19
Operator : RC/JU
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Quant Time: Jul 08 18:19:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM070925.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 08 17:58:12 2025
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 7.869 min Scan# 81
 Delta R.T. 0.006 min
 Lab File: BM050378.D
 Acq: 08 Jul 2025 13:19




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005


Tgt Ion:152 Resp: 332042
 Ion Ratio Lower Upper
 152 100
 150 148.8 126.2 189.4
 115 54.1 45.8 68.8



#2
 1,4-Dioxane
 Concen: 5.278 ng
 RT: 3.334 min Scan# 59
 Delta R.T. 0.000 min
 Lab File: BM050378.D
 Acq: 08 Jul 2025 13:19

Tgt Ion: 88 Resp: 41862
 Ion Ratio Lower Upper
 88 100
 58 77.6 61.0 91.6
 43 31.5 24.6 37.0









#3
 Pyridine
 Concen: 4.981 ng
 RT: 3.734 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: BM050378.D
 Acq: 08 Jul 2025 13:19




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005


Tgt Ion: 79 Resp: 103718
 Ion Ratio Lower Upper
 79 100
 52 76.3 58.3 87.5
 51 42.6 29.6 44.4



#5
 2-Fluorophenol
 Concen: 9.688 ng
 RT: 5.434 min Scan# 416
 Delta R.T. 0.000 min
 Lab File: BM050378.D
 Acq: 08 Jul 2025 13:19

Tgt Ion: 112 Resp: 186878
 Ion Ratio Lower Upper
 112 100
 64 58.1 45.5 68.3
 63 30.4 24.2 36.4









#6
Aniline
Concen: 4.745 ng
RT: 7.187 min Scan# 714
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion: 93 Resp: 148080
Ion Ratio Lower Upper
93 100
66 37.4 30.5 45.7
65 20.0 15.9 23.9



#7
Phenol-d6
Concen: 9.333 ng
RT: 7.016 min Scan# 685
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion: 99 Resp: 226950
Ion Ratio Lower Upper
99 100
42 21.6 15.0 22.6
71 34.5 25.9 38.9









#8
2-Chlorophenol
Concen: 4.632 ng
RT: 7.428 min Scan# 755
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:	128	Resp:	94226
Ion Ratio	Lower	Upper	
128	100		
130	33.0	13.3	53.3
64	47.1	25.6	65.6



#10
Phenol
Concen: 4.778 ng
RT: 7.045 min Scan# 690
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:	94	Resp:	121196
Ion Ratio	Lower	Upper	
94	100		
65	28.0	8.1	48.1
66	35.8	15.8	55.8









#11
bis(2-Chloroethyl)ether
Concen: 4.912 ng
RT: 7.281 min Scan# 711
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion: 93 Resp: 99772
Ion Ratio Lower Upper
93 100
63 79.7 56.9 96.9
95 31.2 12.6 52.6



#12
1,3-Dichlorobenzene
Concen: 5.049 ng
RT: 7.757 min Scan# 811
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:146 Resp: 124875
Ion Ratio Lower Upper
146 100
148 66.9 53.2 79.8
75 29.1 23.9 35.9









#13
1,4-Dichlorobenzene
Concen: 5.103 ng
RT: 7.898 min Scan# 811
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:146 Resp: 128887
Ion Ratio Lower Upper
146 100
148 66.3 51.9 77.9
111 38.2 30.0 45.0



#14
1,2-Dichlorobenzene
Concen: 5.086 ng
RT: 8.216 min Scan# 889
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:146 Resp: 122972
Ion Ratio Lower Upper
146 100
148 63.5 51.5 77.3
111 40.7 32.6 49.0









#16
2,2'-oxybis(1-Chloropropane)
Concen: 5.092 ng
RT: 8.392 min Scan# 919
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion: 45 Resp: 152357
Ion Ratio Lower Upper
45 100
77 13.1 0.0 33.5
79 10.1 0.0 30.3



#17
2-Methylphenol
Concen: 4.655 ng
RT: 8.298 min Scan# 903
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion: 107 Resp: 76571
Ion Ratio Lower Upper
107 100
108 112.0 92.1 138.1
77 45.8 38.2 57.4
79 47.0 36.2 54.4









#18
Hexachloroethane
Concen: 4.840 ng
RT: 8.951 min Scan# 1014
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:117 Resp: 42990
Ion Ratio Lower Upper
117 100
119 99.2 75.6 113.4
201 111.4 87.9 131.9



#19
n-Nitroso-di-n-propylamine
Concen: 4.537 ng
RT: 8.663 min Scan# 965
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion: 70 Resp: 66842
Ion Ratio Lower Upper
70 100
42 66.0 49.7 74.5
101 8.9 7.7 11.5
130 21.1 16.6 24.8









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 10.663 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:136 Resp: 1221582
Ion Ratio Lower Upper
136 100
137 11.1 8.8 13.2
54 8.0 6.3 9.5
68 5.4 4.2 6.2



#22
Acetophenone
Concen: 4.927 ng
RT: 8.681 min Scan# 968
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:105 Resp: 150473
Ion Ratio Lower Upper
105 100
71 2.2 4.4 6.6#
51 29.3 23.0 34.6
120 23.0 18.3 27.5









#23
Nitrobenzene-d5
Concen: 9.234 ng
RT: 9.016 min Scan# 1025
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion: 82 Resp: 221107
Ion Ratio Lower Upper
82 100
128 42.3 35.2 52.8
54 56.7 41.5 62.3



#24
Nitrobenzene
Concen: 4.758 ng
RT: 9.057 min Scan# 1032
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion: 77 Resp: 101661
Ion Ratio Lower Upper
77 100
123 43.7 37.0 55.6
65 14.1 11.0 16.4









#25
Isophorone
Concen: 4.488 ng
RT: 9.586 min Scan# 1122
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion: 82 Resp: 174388
Ion Ratio Lower Upper
82 100
95 7.0 5.6 8.4
138 19.7 15.3 22.9



#26
2-Nitrophenol
Concen: 3.767 ng
RT: 9.769 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion: 139 Resp: 32805
Ion Ratio Lower Upper
139 100
109 31.8 24.6 36.8
65 58.3 39.2 58.8









#27
2,4-Dimethylphenol
Concen: 4.858 ng
RT: 9.828 min Scan# 1163
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:122 Resp: 90013
Ion Ratio Lower Upper
122 100
107 116.8 93.0 139.6
121 56.5 46.1 69.1



#28
bis(2-Chloroethoxy)methane
Concen: 4.822 ng
RT: 10.069 min Scan# 1204
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion: 93 Resp: 124150
Ion Ratio Lower Upper
93 100
95 34.1 25.8 38.6
123 10.8 9.1 13.7





#29

2,4-Dichlorophenol

Concen: 4.438 ng

RT: 10.304 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:162 Resp: 84485

Ion Ratio Lower Upper

162 100

164 65.4 43.7 83.7

98 31.4 13.7 53.7



#30

1,2,4-Trichlorobenzene

Concen: 4.958 ng

RT: 10.522 min Scan# 1281

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19




Tgt Ion:180 Resp: 112916


Ion Ratio Lower Upper

180 100

182 93.2 76.8 115.2

145 28.8 22.6 33.8









#31
Naphthalene
Concen: 5.086 ng
RT: 10.710 min Scan# 11
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:128 Resp: 315550
Ion Ratio Lower Upper
128 100
129 11.5 8.8 13.2
127 12.7 10.4 15.6



#33
4-Chloroaniline
Concen: 4.635 ng
RT: 10.810 min Scan# 1330
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:127 Resp: 121575
Ion Ratio Lower Upper
127 100
129 34.0 25.5 38.3
65 30.4 23.9 35.9
92 18.5 14.6 21.8





#34

Hexachlorobutadiene

Concen: 4.870 ng

RT: 11.010 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:225 Resp: 67689

Ion Ratio Lower Upper

225 100

223 64.5 49.7 74.5

227 66.0 52.1 78.1



#36

4-Chloro-3-methylphenol

Concen: 4.428 ng

RT: 11.927 min Scan# 1520

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19




Tgt Ion:107 Resp: 78716


Ion Ratio Lower Upper

107 100

144 27.7 22.2 33.4

142 84.5 68.7 103.1





#37

2-Methylnaphthalene

Concen: 4.760 ng

RT: 12.322 min Scan# 1

Delta R.T. 0.000 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:142 Resp: 186482

Ion Ratio Lower Upper

142 100

141 90.0 70.8 106.2

115 35.5 27.0 40.4



#38

1-Methylnaphthalene

Concen: 4.796 ng

RT: 12.539 min Scan# 1624

Delta R.T. 0.000 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19




Tgt Ion:142 Resp: 198866


Ion Ratio Lower Upper

142 100

141 91.0 71.9 107.9

116 3.6 3.0 4.6









#39
Acenaphthene-d10
Concen: 20.000 ng
RT: 14.492 min Scan# 1956
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:164 Resp: 768894
Ion Ratio Lower Upper
164 100
162 98.7 79.9 119.9
160 44.3 36.2 54.4



#40
1,2,4,5-Tetrachlorobenzene
Concen: 4.642 ng
RT: 12.686 min Scan# 1649
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:216 Resp: 113496
Ion Ratio Lower Upper
216 100
214 79.3 63.3 94.9
179 21.7 16.4 24.6
108 19.1 13.9 20.9









#42
2,4,6-Tribromophenol
Concen: 7.485 ng
RT: 15.968 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:330 Resp: 69925
Ion Ratio Lower Upper
330 100
332 99.4 76.9 115.3
141 38.9 27.4 41.0



#43
2,4,6-Trichlorophenol
Concen: 4.221 ng
RT: 12.922 min Scan# 1689
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:196 Resp: 63524
Ion Ratio Lower Upper
196 100
198 93.6 75.1 112.7
200 30.8 24.6 37.0









#44
2,4,5-Trichlorophenol
Concen: 4.195 ng
RT: 12.992 min Scan# 1701
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:196	Resp:	68958
Ion Ratio	Lower	Upper
196	100	
198	95.2	77.2 115.8
97	44.6	36.2 54.4
132	22.8	19.9 29.9



#45
2-Fluorobiphenyl
Concen: 9.705 ng
RT: 13.116 min Scan# 1722
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:172	Resp:	604282
Ion Ratio	Lower	Upper
172	100	
171	34.3	27.7 41.5
170	23.7	18.8 28.2









#46
1,1'-Biphenyl
Concen: 4.934 ng
RT: 13.327 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:154 Resp: 281496
Ion Ratio Lower Upper
154 100
153 40.1 19.7 59.7
76 11.9 0.0 30.9



#47
2-Chloronaphthalene
Concen: 4.918 ng
RT: 13.369 min Scan# 1765
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:162 Resp: 223672
Ion Ratio Lower Upper
162 100
127 35.5 28.3 42.5
164 32.6 26.6 40.0





#48

2-Nitroaniline

Concen: 3.738 ng

RT: 13.563 min Scan# 1798

Delta R.T. 0.000 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion: 65 Resp: 37953

Ion Ratio Lower Upper

65 100

92 68.3 54.3 81.5

138 104.7 87.8 131.6



#49

Acenaphthylene

Concen: 4.601 ng

RT: 14.216 min Scan# 1909

Delta R.T. 0.000 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19




Tgt Ion:152 Resp: 316292


Ion Ratio Lower Upper

152 100

151 19.7 16.3 24.5

153 12.4 10.6 15.8





#50

Dimethylphthalate

Concen: 4.805 ng

RT: 13.945 min Scan# 1863

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:163 Resp: 254362

Ion Ratio Lower Upper

163 100

194 4.2 3.4 5.2

164 9.9 8.2 12.2



#51

2,6-Dinitrotoluene

Concen: 3.710 ng

RT: 14.057 min Scan# 1882

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19




Tgt Ion:165 Resp: 37711


Ion Ratio Lower Upper

165 100

63 61.5 42.4 63.6

89 60.8 42.2 63.2









#52
Acenaphthene
Concen: 4.779 ng
RT: 14.557 min Scan# 1967
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:154 Resp: 208840
Ion Ratio Lower Upper
154 100
153 114.3 89.2 133.8
152 56.4 42.6 64.0



#53
3-Nitroaniline
Concen: 3.628 ng
RT: 14.380 min Scan# 1937
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:138 Resp: 39220
Ion Ratio Lower Upper
138 100
108 16.1 8.8 13.2#
92 121.7 85.0 127.6









#55
Dibenzenofuran
Concen: 4.988 ng
RT: 14.886 min Scan# 2015
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:168 Resp: 335867
Ion Ratio Lower Upper
168 100
139 35.0 27.7 41.5
169 13.4 10.7 16.1



#57
2,4-Dinitrotoluene
Concen: 5.513 ng
RT: 14.839 min Scan# 2015
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:165 Resp: 44983
Ion Ratio Lower Upper
165 100
63 50.1 32.5 48.7#
89 77.4 54.0 81.0
182 3.3 2.0 3.0#





#58

Fluorene

Concen: 4.671 ng

RT: 15.533 min Scan# 2134

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:166 Resp: 259083

Ion Ratio Lower Upper

166 100

165 98.7 77.0 115.6

167 13.2 10.9 16.3



#59

2,3,4,6-Tetrachlorophenol

Concen: 4.158 ng

RT: 15.110 min Scan# 2061

Delta R.T. 0.000 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Tgt Ion:232 Resp: 57532




Ion Ratio Lower Upper


232 100

131 46.5 33.4 50.0

130 2.5 1.8 2.6

166 31.3 24.2 36.4





#60

Diethylphthalate

Concen: 4.625 ng

RT: 15.310 min Scan# 2096

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:149 Resp: 233863

Ion Ratio Lower Upper

149 100

177 22.2 18.8 28.2

150 11.9 9.8 14.8



#61

4-Chlorophenyl-phenylether

Concen: 4.557 ng

RT: 15.533 min Scan# 2133

Delta R.T. 0.000 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19




Tgt Ion:204 Resp: 132153


Ion Ratio Lower Upper

204 100

206 32.7 26.6 39.8

141 57.0 39.9 59.9









#62
4-Nitroaniline
Concen: 5.398 ng
RT: 15.533 min Scan# 2133
Delta R.T. -0.012 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:138 Resp: 36922
Ion Ratio Lower Upper
138 100
92 55.5 24.5 64.5
108 91.3 52.5 92.5



#63
Azobenzene
Concen: 4.490 ng
RT: 15.821 min Scan# 2182
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion: 77 Resp: 208493
Ion Ratio Lower Upper
77 100
182 26.8 10.0 50.0
105 20.6 2.0 42.0
51 33.3 11.7 51.7





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 17.227 min Scan# 2421

Delta R.T. 0.000 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:188 Resp: 1453955

Ion Ratio Lower Upper

188 100

94 7.5 6.0 9.0

80 8.3 6.5 9.7



#66

n-Nitrosodiphenylamine

Concen: 4.566 ng

RT: 15.739 min Scan# 2168

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19




Tgt Ion:169 Resp: 207747


Ion Ratio Lower Upper

169 100

168 68.9 53.2 79.8

167 36.8 28.7 43.1









#67
4-Bromophenyl-phenylether
Concen: 4.450 ng
RT: 16.421 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:248 Resp: 71279
Ion Ratio Lower Upper
248 100
250 93.3 77.9 116.9
141 73.2 49.7 74.5



#68
Hexachlorobenzene
Concen: 4.587 ng
RT: 16.539 min Scan# 2304
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:284 Resp: 86781
Ion Ratio Lower Upper
284 100
142 34.7 26.8 40.2
249 31.0 24.5 36.7









#69
 Atrazine
 Concen: 4.046 ng
 RT: 16.686 min Scan# 21
 Delta R.T. -0.006 min
 Lab File: BM050378.D
 Acq: 08 Jul 2025 13:19




Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005


Tgt Ion:200 Resp: 60444
 Ion Ratio Lower Upper
 200 100
 173 26.0 5.6 45.6
 215 47.0 29.6 69.6



#71
 Phenanthrene
 Concen: 4.943 ng
 RT: 17.268 min Scan# 2428
 Delta R.T. 0.000 min
 Lab File: BM050378.D
 Acq: 08 Jul 2025 13:19

Tgt Ion:178 Resp: 406647
 Ion Ratio Lower Upper
 178 100
 176 19.1 15.4 23.0
 179 14.7 12.5 18.7





#72

Anthracene

Concen: 4.563 ng

RT: 17.357 min Scan# 2443

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:178 Resp: 373674

Ion Ratio Lower Upper

178 100

176 18.2 14.9 22.3

179 14.6 12.2 18.4



#73

Carbazole

Concen: 4.605 ng

RT: 17.621 min Scan# 2488

Delta R.T. 0.000 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19




Tgt Ion:167 Resp: 341196


Ion Ratio Lower Upper

167 100

166 21.7 17.2 25.8

139 12.1 9.9 14.9









#74
Di-n-butylphthalate
Concen: 4.274 ng
RT: 18.198 min Scan# 21
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:149 Resp: 346654
Ion Ratio Lower Upper
149 100
150 9.4 7.3 10.9
104 5.7 3.9 5.9



#75
Fluoranthene
Concen: 4.421 ng
RT: 19.274 min Scan# 2769
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:202 Resp: 395390
Ion Ratio Lower Upper
202 100
101 9.6 0.0 28.3
203 17.6 0.0 37.2





#76

Chrysene-d12

Concen: 20.000 ng

RT: 21.450 min Scan# 31

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:240 Resp: 1431554

Ion Ratio Lower Upper

240 100

120 9.6 6.7 10.1

236 26.3 20.7 31.1



#78

Pyrene

Concen: 4.676 ng

RT: 19.633 min Scan# 2830

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19




Tgt Ion:202 Resp: 428758


Ion Ratio Lower Upper

202 100

200 20.9 16.9 25.3

203 17.7 14.1 21.1





#79

Terphenyl-d14

Concen: 9.630 ng

RT: 19.839 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:244 Resp: 783562

Ion Ratio Lower Upper

244 100

212 6.9 5.7 8.5

122 9.5 6.2 9.2#



#80

Butylbenzylphthalate

Concen: 3.757 ng

RT: 20.533 min Scan# 2983

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19




Tgt Ion:149 Resp: 113580


Ion Ratio Lower Upper

149 100

91 68.6 51.6 77.4

206 19.0 17.8 26.8





#81

Benzo(a)anthracene

Concen: 4.620 ng

RT: 21.433 min Scan# 3137

Delta R.T. -0.006 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:228 Resp: 430980

Ion Ratio Lower Upper

228 100

226 27.6 21.8 32.8

229 20.1 16.0 24.0



#83

Chrysene

Concen: 4.727 ng

RT: 21.497 min Scan# 3147

Delta R.T. 0.000 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19




Tgt Ion:228 Resp: 415885


Ion Ratio Lower Upper

228 100

226 29.0 24.2 36.4

229 19.8 15.8 23.6









#84
Bis(2-ethylhexyl)phthalate
Concen: 3.724 ng
RT: 21.368 min Scan# 3125
Delta R.T. 0.000 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:149 Resp: 178701
Ion Ratio Lower Upper
149 100
167 27.1 22.7 34.1
279 3.9 3.5 5.3



#86
Perylene-d12
Concen: 20.000 ng
RT: 24.491 min Scan# 3656
Delta R.T. -0.006 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:264 Resp: 1410719
Ion Ratio Lower Upper
264 100
260 25.0 19.2 28.8
265 21.4 17.8 26.6





#87

Indeno(1,2,3-cd)pyrene

Concen: 4.249 ng

RT: 27.921 min Scan# 41

Delta R.T. -0.018 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:276 Resp: 426152

Ion Ratio Lower Upper

276 100

138 17.1 17.0 25.6

277 25.5 20.0 30.0



#88

Benzo(b)fluoranthene

Concen: 4.431 ng

RT: 23.527 min Scan# 3492

Delta R.T. -0.012 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19




Tgt Ion:252 Resp: 384470


Ion Ratio Lower Upper

252 100

253 21.6 17.8 26.6

125 9.4 6.3 9.5









#89
Benzo(k)fluoranthene
Concen: 4.330 ng
RT: 23.591 min Scan# 31
Delta R.T. -0.012 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19




Instrument :
BNA_M
ClientSampleId :
SSTDICC005


Tgt Ion:252 Resp: 389878
Ion Ratio Lower Upper
252 100
253 22.9 17.6 26.4
125 9.3 6.2 9.2#



#90
Benzo(a)pyrene
Concen: 4.177 ng
RT: 24.344 min Scan# 3631
Delta R.T. -0.012 min
Lab File: BM050378.D
Acq: 08 Jul 2025 13:19

Tgt Ion:252 Resp: 339308
Ion Ratio Lower Upper
252 100
253 20.7 17.8 26.6
125 9.9 7.0 10.4





#91

Dibenzo(a,h)anthracene

Concen: 4.269 ng

RT: 27.991 min Scan# 41

Delta R.T. -0.012 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Instrument :

BNA_M

ClientSampleId :

SSTDICC005





Tgt Ion:278 Resp: 360576

Ion Ratio Lower Upper

278 100

139 13.5 10.8 16.2

279 23.1 19.4 29.0



#92

Benzo(g,h,i)perylene

Concen: 4.382 ng

RT: 28.985 min Scan# 4420

Delta R.T. -0.023 min

Lab File: BM050378.D

Acq: 08 Jul 2025 13:19

Tgt Ion:276 Resp: 349804

Ion Ratio Lower Upper

276 100

277 24.4 19.1 28.7

138 20.8 14.9 22.3

