

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF061925

Review By	Rahul	Review On	6/23/2025 3:55:52 PM
Supervise By	Jagrut	Supervise On	6/24/2025 3:17:49 PM
SubDirectory	BF061925	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12670,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF142787.D	19 Jun 2025 16:37		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF142788.D	19 Jun 2025 17:07		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF142789.D	19 Jun 2025 17:38		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF142790.D	19 Jun 2025 18:08		RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF142791.D	19 Jun 2025 18:39		RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF142792.D	19 Jun 2025 19:09		RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF142793.D	19 Jun 2025 19:40		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF142794.D	19 Jun 2025 20:10		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF142795.D	19 Jun 2025 20:40	Compound #09 removed from 80 PPM.	RC/JU	Ok
10	SSTDICV040	ICVBF061925	BF142796.D	19 Jun 2025 21:10		RC/JU	Ok
11	PB168536BL	PB168536BL	BF142797.D	19 Jun 2025 22:10		RC/JU	Ok
12	PB168536BS	PB168536BS	BF142798.D	19 Jun 2025 22:40		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070925

Review By	Rahul	Review On	7/9/2025 3:55:15 PM
Supervise By	Jagrut	Supervise On	7/10/2025 5:36:41 PM
SubDirectory	BF070925	HP Acquire Method	BNA_F
		HP Processing Method	bf061925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF143047.D	09 Jul 2025 10:22		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF143048.D	09 Jul 2025 10:52		RC/JU	Ok,M
3	PB168737BL	PB168737BL	BF143049.D	09 Jul 2025 11:23		RC/JU	Ok
4	PB168737BS	PB168737BS	BF143050.D	09 Jul 2025 11:54		RC/JU	Ok,M
5	PB168750BL	PB168750BL	BF143051.D	09 Jul 2025 12:24		RC/JU	Ok
6	PB168750BS	PB168750BS	BF143052.D	09 Jul 2025 12:54		RC/JU	Ok,M
7	Q2489-06	G2(6-12)	BF143053.D	09 Jul 2025 13:27		RC/JU	Ok
8	Q2489-07	G1(0-6)	BF143054.D	09 Jul 2025 13:57		RC/JU	Ok
9	Q2489-08	G1(6-12)	BF143055.D	09 Jul 2025 14:28		RC/JU	Ok
10	Q2489-03	G3(0-6)	BF143056.D	09 Jul 2025 14:58		RC/JU	Ok
11	Q2489-04	G3(6-12)	BF143057.D	09 Jul 2025 15:28		RC/JU	Ok
12	Q2489-05	G2(0-6)	BF143058.D	09 Jul 2025 15:59		RC/JU	Ok
13	Q2517-01	TP-14	BF143059.D	09 Jul 2025 16:29		RC/JU	Ok
14	Q2517-01MS	TP-14MS	BF143060.D	09 Jul 2025 17:00		RC/JU	Ok,M
15	Q2517-01MSD	TP-14MSD	BF143061.D	09 Jul 2025 17:30		RC/JU	Ok,M
16	Q2529-07	TP-101	BF143062.D	09 Jul 2025 18:01		RC/JU	Ok
17	Q2487-13	G2(0-6)	BF143063.D	09 Jul 2025 18:31	Need further 2X Dilution	RC/JU	Dilution
18	Q2487-09	G4(0-6)	BF143064.D	09 Jul 2025 19:02		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF070925

Review By	Rahul	Review On	7/9/2025 3:55:15 PM		
Supervise By	Jagrut	Supervise On	7/10/2025 5:36:41 PM		
SubDirectory	BF070925	HP Acquire Method	BNA_F	HP Processing Method	bf061925
STD. NAME	STD REF.#				
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840				
CCC	SP6836				
Internal Standard/PEM	S12673,10ul/1000ul sample				
ICV/I.BLK	SP6770				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	Q2529-06	TP-102	BF143065.D	09 Jul 2025 19:33		RC/JU	Ok
20	Q2527-01	ARS20-0015	BF143066.D	09 Jul 2025 20:04	Internal standard fail	RC/JU	Ok,M
21	Q2487-16	G1(6-12)	BF143067.D	09 Jul 2025 20:34	Internal standard fail	RC/JU	Ok,M
22	Q2487-11	G3(0-6)	BF143068.D	09 Jul 2025 21:04	Internal standard fail	RC/JU	Ok,M
23	Q2487-15	G1(0-6)	BF143069.D	09 Jul 2025 21:35	Internal standard fail	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM070925

Review By	Rahul	Review On	7/9/2025 9:48:51 AM
Supervise By	Jagrut	Supervise On	7/9/2025 2:15:39 PM
SubDirectory	BM070925	HP Acquire Method	BNA_M
		HP Processing Method	bm070925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050376.D	08 Jul 2025 11:59		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM050377.D	08 Jul 2025 12:39		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM050378.D	08 Jul 2025 13:19		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BM050379.D	08 Jul 2025 14:00		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM050380.D	08 Jul 2025 14:40		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BM050381.D	08 Jul 2025 15:20	Compound#32,54,57,62,65 Kept on LR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BM050382.D	08 Jul 2025 16:01		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BM050383.D	08 Jul 2025 16:41		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BM050384.D	08 Jul 2025 17:22	Compound#09 removed from 80 ppm	RC/JU	Ok
10	SSTDICV040	ICVBM070925	BM050385.D	08 Jul 2025 18:05	Comp#77 Failed from High side	RC/JU	Ok,M
11	PB168722BL	PB168722BL	BM050386.D	08 Jul 2025 19:26		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM071025

Review By	Rahul	Review On	7/11/2025 12:08:59 PM
Supervise By	Jagrut	Supervise On	7/11/2025 2:56:38 PM
SubDirectory	BM071025	HP Acquire Method	BNA_M
		HP Processing Method	bm070925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050387.D	10 Jul 2025 13:13		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM050388.D	10 Jul 2025 13:53		RC/JU	Ok,M
3	PB168689TB	PB168689TB	BM050389.D	10 Jul 2025 14:33		RC/JU	Ok
4	Q2529-03	TP-79	BM050390.D	10 Jul 2025 15:17		RC/JU	Ok
5	Q2529-05	TP-98	BM050391.D	10 Jul 2025 15:57		RC/JU	Ok
6	Q2529-08	TP-89	BM050392.D	10 Jul 2025 16:37		RC/JU	Ok
7	Q2529-09	TP-33	BM050393.D	10 Jul 2025 17:18		RC/JU	Ok
8	Q2543-02	TP-49	BM050394.D	10 Jul 2025 17:58		RC/JU	Ok,M
9	Q2543-02MS	TP-49MS	BM050395.D	10 Jul 2025 18:38		RC/JU	Ok,M
10	Q2543-02MSD	TP-49MSD	BM050396.D	10 Jul 2025 19:18		RC/JU	Ok,M
11	Q2489-01	G4(0-6)	BM050397.D	10 Jul 2025 19:59		RC/JU	Ok
12	Q2489-01MS	G4(0-6)MS	BM050398.D	10 Jul 2025 20:39		RC/JU	Ok,M
13	Q2489-01MSD	G4(0-6)MSD	BM050399.D	10 Jul 2025 21:19		RC/JU	Ok,M
14	Q2489-02	G4(6-12)	BM050400.D	10 Jul 2025 21:59		RC/JU	Ok
15	Q2489-03	G3(0-6)	BM050401.D	10 Jul 2025 22:39		RC/JU	Ok
16	Q2543-03	TP-23	BM050402.D	10 Jul 2025 23:19		RC/JU	Ok
17	Q2543-04	TP-23-99	BM050403.D	10 Jul 2025 23:59		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP070325

Review By	Rahul	Review On	7/7/2025 11:36:11 AM
Supervise By	Jagrut	Supervise On	7/7/2025 12:15:03 PM
SubDirectory	BP070325	HP Acquire Method	BNA_P
		HP Processing Method	bp070325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12672,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025070.D	03 Jul 2025 08:58		CG/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP025071.D	03 Jul 2025 09:39		CG/JU	Ok
3	SSTDICC005	SSTDICC005	BP025072.D	03 Jul 2025 10:21		CG/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP025073.D	03 Jul 2025 11:02		CG/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP025074.D	03 Jul 2025 11:44		CG/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP025075.D	03 Jul 2025 12:25	compound #77 failed in the calibration.	CG/JU	Ok,M
7	SSTDICC050	SSTDICC050	BP025076.D	03 Jul 2025 13:07	Compound#26,32,48,54,57,65,80,85 Kept on LR	CG/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP025077.D	03 Jul 2025 13:49		CG/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP025078.D	03 Jul 2025 14:31	Compound#09 removed from 80 ppm	CG/JU	Ok,M
10	SSTDICV040	ICVBP070325	BP025079.D	03 Jul 2025 16:18		CG/JU	Ok,M
11	PB168694BL	PB168694BL	BP025080.D	03 Jul 2025 16:59	Analyzed for contamination check	CG/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP071025

Review By	Rahul	Review On	7/11/2025 12:10:50 PM
Supervise By	Jagrut	Supervise On	7/11/2025 2:55:32 PM
SubDirectory	BP071025	HP Acquire Method	BNA_P
		HP Processing Method	bp070325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025081.D	10 Jul 2025 13:18		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025082.D	10 Jul 2025 14:04	CCC fail high side for com. #77	CG/JU	Ok,M
3	PB168767BL	PB168767BL	BP025083.D	10 Jul 2025 14:48		CG/JU	Ok
4	PB168767BS	PB168767BS	BP025084.D	10 Jul 2025 15:30		CG/JU	Ok,M
5	Q2527-03	LAW-PAD-070825	BP025085.D	10 Jul 2025 16:15		CG/JU	Ok
6	Q2524-01	OLD RETENTION BAS	BP025086.D	10 Jul 2025 16:57		CG/JU	Ok
7	Q2523-03	PAD-070825-2	BP025087.D	10 Jul 2025 17:39		CG/JU	Ok
8	Q2529-10	TP-30	BP025088.D	10 Jul 2025 18:21		CG/JU	Ok
9	Q2529-04	TP-95	BP025089.D	10 Jul 2025 19:03		CG/JU	Ok
10	Q2529-01	TP-91	BP025090.D	10 Jul 2025 19:44		CG/JU	Ok,M
11	Q2519-01	TP-15	BP025091.D	10 Jul 2025 20:26		CG/JU	Ok
12	Q2523-01	MOO-25-0190	BP025092.D	10 Jul 2025 21:08	Internal standard fail	CG/JU	Ok,M
13	Q2523-02	PAD-070825-1	BP025093.D	10 Jul 2025 21:50	Internal standard fail	CG/JU	Ok
14	Q2526-03	LAW-25-0102	BP025094.D	10 Jul 2025 22:31		CG/JU	Ok
15	Q2487-13DL	G2(0-6)DL	BP025095.D	10 Jul 2025 23:13	Need further 2X Dilution	CG/JU	Dilution
16	Q2526-01	LAW-25-0101	BP025096.D	10 Jul 2025 23:55		CG/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP071625

Review By	Review On
Supervise By	Supervise On
SubDirectory	BP071625
HP Acquire Method	BNA_P
HP Processing Method	bp070325
STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025154.D	16 Jul 2025 11:19		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025155.D	16 Jul 2025 12:42	CCC fail high side for com. #26,32,41,54,65,77,82 and #84 and surrogate #42 is high	CG/JU	Ok,NR
3	PB168832BL	PB168832BL	BP025156.D	16 Jul 2025 13:23		CG/JU	Ok
4	PB168832BS	PB168832BS	BP025157.D	16 Jul 2025 14:05		CG/JU	Ok,NR
5	Q2586-01	TP-16	BP025158.D	16 Jul 2025 14:49		CG/JU	Ok
6	Q2586-01MS	TP-16MS	BP025159.D	16 Jul 2025 15:31		CG/JU	Ok,NR
7	Q2586-01MSD	TP-16MSD	BP025160.D	16 Jul 2025 16:12		CG/JU	Ok,NR
8	Q2600-05	STOCK-PILE	BP025161.D	16 Jul 2025 16:54	CCC fail high side for com. #84 (Below CRQL)	CG/JU	Ok
9	Q2529-02	TP-80	BP025162.D	16 Jul 2025 17:36		CG/JU	Ok,NR
10	Q2600-01	TRENCH	BP025163.D	16 Jul 2025 18:17	CCC Fail High side for comp #84 (Below CRQL)	CG/JU	Ok,NR
11	Q2600-09	END-OF-TRENCH	BP025164.D	16 Jul 2025 18:59	CCC Fail High side for comp #84	CG/JU	ReRun
12	Q2598-01	OK-01-071425	BP025165.D	16 Jul 2025 19:40	Need 2X Dilution	CG/JU	Dilution
13	Q2592-01	WC-SOIL-20250711	BP025166.D	16 Jul 2025 20:22	CCC Fail High side for comp #32(Below CRQL),84	CG/JU	ReRun
14	Q2589-01	AU-06-071125	BP025167.D	16 Jul 2025 21:04	Need Straight Run	CG/JU	Not Ok
15	Q2565-04	MOO-25-0196	BP025168.D	16 Jul 2025 21:45	CCC Fail High side for comp #84	CG/JU	ReRun
16	Q2611-01	EO-02-071525	BP025169.D	16 Jul 2025 22:26		CG/JU	Ok,NR

M : Manual Integration