

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2529

Client ID : SSTDCCC040

Date Analyzed: 07/16/2025

Lab File ID: BP025155.D

Time Analyzed: 12:42

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	564997	7.443	2237580	10.20	1465070	14.10
UPPER LIMIT	1129990	7.943	4475160	10.696	2930140	14.595
LOWER LIMIT	282499	6.943	1118790	9.696	732535	13.595
EPA SAMPLE NO.						
01 TP-80	576516	7.44	2197510	10.19	1357110	14.10

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.