

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: CHEM

SDG NO.: Q2529

Client ID: SSTDICCC040

Date Analyzed: 07/08/2025

Lab File ID: BM050381.D

Time Analyzed: 15:20

Instrument ID: BNA_M

GC Column: ZB-GR ID: 0.25 (mm)

		IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
	12 HOUR STD	1834050	17.227	1904000	21.456	1842290	24.497
	UPPER LIMIT	3668100	17.727	3808000	21.956	3684580	24.997
	LOWER LIMIT	917025	16.727	952000	20.956	921145	23.997
	EPA SAMPLE NO.						
01	TP-79	2509640	17.22	2687320	21.45	2766870	24.49
02	TP-98	2502020	17.23	2656850	21.45	2652820	24.49
03	TP-89	2477430	17.23	2462750	21.45	2316010	24.49
04	TP-33	2414090	17.22	2506900	21.45	2387080	24.49

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.