

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BP025070.D
Instrument ID: BNA_P

Contract: CAMP02
SDG NO.: Q2529
DFTPP Injection Date: 07/03/2025
DFTPP Injection Time: 08:58

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.7) 1
69	Mass 69 relative abundance	31.4
70	Less than 2.0% of mass 69	0.1 (0.4) 1
197	Less than 2.0% of mass 198	0.3
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	14.3
442	Greater than 50% of mass 198	92.8
443	15.0 - 24.0% of mass 442	18 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC2.5	SSTDICC2.5	BP025071.D	07/03/2025	09:39
SSTDICC005	SSTDICC005	BP025072.D	07/03/2025	10:21
SSTDICC010	SSTDICC010	BP025073.D	07/03/2025	11:02
SSTDICC020	SSTDICC020	BP025074.D	07/03/2025	11:44
SSTDICCC040	SSTDICCC040	BP025075.D	07/03/2025	12:25
SSTDICC050	SSTDICC050	BP025076.D	07/03/2025	13:07
SSTDICC060	SSTDICC060	BP025077.D	07/03/2025	13:49
SSTDICC080	SSTDICC080	BP025078.D	07/03/2025	14:31