

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: CAMP02
 Lab Code: ACE SDG No.: Q2529
 Instrument ID: MSVOA_W Calibration Date/Time: 07/10/2025 08:47
 Lab File ID: VW031791.D Init. Calib. Date(s): 06/30/2025 06/30/2025
 Heated Purge: (Y/N) Y Init. Calib. Time(s): 09:54 12:55
 GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Methylene Chloride	50.000	59.33	-18.666	20