

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2529</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>CDM Smith</u>	Datafile :	<u>VW031769.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VW0709SBS01	Dichlorodifluoromethane	20	23.9	ug/Kg	119			64	136	
	Chloromethane	20	23.4	ug/Kg	117			52	151	
	Vinyl chloride	20	24.1	ug/Kg	121			56	148	
	Bromomethane	20	23.2	ug/Kg	116			58	141	
	Chloroethane	20	23.2	ug/Kg	116			69	130	
	Trichlorofluoromethane	20	20.8	ug/Kg	104			69	134	
	1,1,2-Trichlorotrifluoroethane	20	23.9	ug/Kg	119			81	123	
	1,1-Dichloroethene	20	23.8	ug/Kg	119			79	121	
	Acetone	100	130	ug/Kg	130			40	171	
	Carbon disulfide	20	22.8	ug/Kg	114			59	130	
	Methyl tert-butyl Ether	20	24.4	ug/Kg	122			77	129	
	Methyl Acetate	20	23.1	ug/Kg	116			69	149	
	Methylene Chloride	20	35.2	ug/Kg	176		*	72	131	
	trans-1,2-Dichloroethene	20	24.2	ug/Kg	121			80	123	
	1,1-Dichloroethane	20	24.1	ug/Kg	121			82	123	
	Cyclohexane	20	23.8	ug/Kg	119			76	122	
	2-Butanone	100	130	ug/Kg	130			69	131	
	Carbon Tetrachloride	20	22.9	ug/Kg	115			76	129	
	cis-1,2-Dichloroethene	20	24.8	ug/Kg	124		*	82	123	
	Bromochloromethane	20	24.3	ug/Kg	121			80	127	
	Chloroform	20	24.3	ug/Kg	121			82	125	
	1,1,1-Trichloroethane	20	24.6	ug/Kg	123			80	126	
	Methylcyclohexane	20	22.9	ug/Kg	115			77	123	
	Benzene	20	23.9	ug/Kg	119			84	121	
	1,2-Dichloroethane	20	23.4	ug/Kg	117			81	126	
	Trichloroethene	20	22.7	ug/Kg	114			83	122	
	1,2-Dichloropropane	20	23.5	ug/Kg	117			83	122	
	Bromodichloromethane	20	23.3	ug/Kg	117			82	123	
	4-Methyl-2-Pentanone	100	120	ug/Kg	120			70	135	
	Toluene	20	24.1	ug/Kg	121			83	122	
	t-1,3-Dichloropropene	20	23.4	ug/Kg	117			78	124	
	cis-1,3-Dichloropropene	20	23.5	ug/Kg	117			81	122	
	1,1,2-Trichloroethane	20	23.7	ug/Kg	119			82	125	
	2-Hexanone	100	120	ug/Kg	120			66	138	
	Dibromochloromethane	20	22.3	ug/Kg	112			79	125	
	1,2-Dibromoethane	20	23.8	ug/Kg	119			80	125	
	Tetrachloroethene	20	23.7	ug/Kg	119			83	125	
	Chlorobenzene	20	23.8	ug/Kg	119			84	122	
	Ethyl Benzene	20	23.1	ug/Kg	116			82	124	
	m/p-Xylenes	40	46.8	ug/Kg	117			83	124	
	o-Xylene	20	23.7	ug/Kg	119			83	123	
	Styrene	20	23.9	ug/Kg	119			82	124	
	Bromoform	20	23.7	ug/Kg	119			75	127	
	Isopropylbenzene	20	22.7	ug/Kg	114			82	124	
	1,1,2,2-Tetrachloroethane	20	23.3	ug/Kg	117			77	127	
	1,3-Dichlorobenzene	20	24.1	ug/Kg	121			83	122	
	1,4-Dichlorobenzene	20	23.4	ug/Kg	117			84	121	
	1,2-Dichlorobenzene	20	24.3	ug/Kg	121			83	124	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2529</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>CDM Smith</u>	Datafile :	<u>VW031769.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VW0709SBS01	1,2-Dibromo-3-Chloropropane	20	21.9	ug/Kg	110			66	134	
	1,2,4-Trichlorobenzene	20	24.2	ug/Kg	121			78	127	
	1,2,3-Trichlorobenzene	20	23.4	ug/Kg	117			70	137	