

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2532

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4100	4000	103	95 - 105	P	07/10/2025	18:39	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2532

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	118	100	118	80 - 120	P	07/10/2025	18:51	LB136434

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2532

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5140	5000	103	90 - 110	P	07/10/2025	19:43	LB136434
CCV02	Iron	5010	5000	100	90 - 110	P	07/10/2025	20:59	LB136434
CCV03	Iron	5140	5000	103	90 - 110	P	07/10/2025	22:32	LB136434
CCV04	Iron	4890	5000	98	90 - 110	P	07/10/2025	23:29	LB136434
CCV05	Iron	4950	5000	99	90 - 110	P	07/11/2025	00:47	LB136434
CCV06	Iron	4890	5000	98	90 - 110	P	07/11/2025	01:47	LB136434
CCV07	Iron	4630	5000	93	90 - 110	P	07/11/2025	03:17	LB136434
CCV08	Iron	4780	5000	96	90 - 110	P	07/11/2025	04:18	LB136434
CCV09	Iron	4810	5000	96	90 - 110	P	07/11/2025	05:19	LB136434
CCV10	Iron	4700	5000	94	90 - 110	P	07/11/2025	06:16	LB136434
CCV11	Iron	4700	5000	94	90 - 110	P	07/11/2025	06:43	LB136434



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2532

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Iron	4090	4000	102	95 - 105	P	07/15/2025	19:27	LB136488



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2532

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Iron	118	100	118	80 - 120	P	07/15/2025	19:37	LB136488

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q2532

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Iron	5250	5000	105	90 - 110	P	07/15/2025	20:25	LB136488
CCV02	Iron	5230	5000	104	90 - 110	P	07/15/2025	22:12	LB136488
CCV03	Iron	5090	5000	102	90 - 110	P	07/15/2025	22:39	LB136488
