

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: Tully Environmental, Inc

SDG No.: Q2532

Contract: TULL01

Lab Code: ACE

ICS Source: EPA

Instrument ID: P5

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Iron	104000	100000	104	85600	116500	07/10/2025	19:11	LB136434
ICSAB01	Iron	98700	99000	100	84400	114500	07/10/2025	19:15	LB136434
ICSA	Iron	112000	100000	112	85600	116500	07/10/2025	19:20	LB136434
ICSAB	Iron	111000	99000	112	84400	114500	07/10/2025	19:24	LB136434
ICSA01	Iron	99100	100000	99	85600	116500	07/15/2025	20:01	LB136488
ICSAB01	Iron	99500	99000	100	84400	114500	07/15/2025	20:05	LB136488
ICSA	Iron	111000	100000	111	85600	116500	07/15/2025	20:10	LB136488
ICSAB	Iron	104000	99000	105	84400	114500	07/15/2025	20:20	LB136488