

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q2533

Contract: TULL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB168772BS							
Copper	ug/L	150	164		109	85 - 115	P
Iron	ug/L	1500	1720		115	85 - 115	P
Lead	ug/L	500	464		93	85 - 115	P