

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2533</u>	Analytical Method:	<u>SW624.1</u>
Client:	<u>Tully Environmental, Inc</u>	Datafile :	<u>VN087308.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0709WBSD01	Benzene	20	17.1	ug/L	86	0		65	135	20
	Toluene	20	18.0	ug/L	90	2		70	130	20
	Ethyl Benzene	20	17.7	ug/L	89	5		60	140	20
	m/p-Xylenes	40	38.6	ug/L	97	7		87	111	20
	o-Xylene	20	18.4	ug/L	92	2		87	111	20