

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: TULL01
 Lab Code: ACE SDG NO.: Q2533
 Lab File ID: VN087306.D Date Analyzed: 07/09/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:57
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	43430	7.82	217232	9.11	209262	11.87
UPPER LIMIT	86860	8.318	434464	9.606	418524	12.365
LOWER LIMIT	21715	7.318	108616	8.606	104631	11.365
EPA SAMPLE NO.						
001 WILLETS PT BLVD (JUNE)	40947	7.82	205920	9.11	191589	11.87
002-35th-AVE (JUNE)	39994	7.82	207389	9.11	187985	11.87
VN0709WBL01	41807	7.82	215919	9.11	195049	11.87
VN0709WBS01	42206	7.82	198713	9.11	190461	11.87
VN0709WBSD01	45337	7.82	235043	9.11	219225	11.87

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.