



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2543

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB25	Mercury	0.076	+/-0.2	U	0.20	CV	07/11/2025	14:15	LB136438

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2543

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB82	Mercury	0.076	+/-0.2	U	0.20	CV	07/11/2025	14:25	LB136438
CCB83	Mercury	0.076	+/-0.2	U	0.20	CV	07/11/2025	15:01	LB136438
CCB84	Mercury	0.076	+/-0.2	U	0.20	CV	07/11/2025	15:29	LB136438
CCB85	Mercury	0.076	+/-0.2	U	0.20	CV	07/11/2025	16:04	LB136438
CCB86	Mercury	0.076	+/-0.2	U	0.20	CV	07/11/2025	16:34	LB136438

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2543

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	11.3	+/-50	U	100	P	07/10/2025	18:05	LB136428
	Antimony	6.76	+/-25	U	50.0	P	07/10/2025	18:05	LB136428
	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	18:05	LB136428
	Barium	14.6	+/-50	U	100	P	07/10/2025	18:05	LB136428
	Beryllium	0.56	+/-3	U	6.00	P	07/10/2025	18:05	LB136428
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	18:05	LB136428
	Calcium	234	+/-1000	U	2000	P	07/10/2025	18:05	LB136428
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	18:05	LB136428
	Cobalt	2.26	+/-15	U	30.0	P	07/10/2025	18:05	LB136428
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	18:05	LB136428
	Iron	23.4	+/-50	U	100	P	07/10/2025	18:05	LB136428
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	18:05	LB136428
	Magnesium	244	+/-1000	U	2000	P	07/10/2025	18:05	LB136428
	Manganese	5.94	+/-10	U	20.0	P	07/10/2025	18:05	LB136428
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	18:05	LB136428
	Potassium	918	+/-1000	U	2000	P	07/10/2025	18:05	LB136428
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	18:05	LB136428
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	18:05	LB136428
	Sodium	868	+/-1000	U	2000	P	07/10/2025	18:05	LB136428
	Thallium	4.38	+/-20	U	40.0	P	07/10/2025	18:05	LB136428
	Vanadium	6.26	+/-20	U	40.0	P	07/10/2025	18:05	LB136428
	Zinc	16.7	+/-20	U	40.0	P	07/10/2025	18:05	LB136428

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2543

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	11.3	+/-50	U	100	P	07/10/2025	18:44	LB136428
	Antimony	6.76	+/-25	U	50.0	P	07/10/2025	18:44	LB136428
	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	18:44	LB136428
	Barium	14.6	+/-50	U	100	P	07/10/2025	18:44	LB136428
	Beryllium	0.56	+/-3	U	6.00	P	07/10/2025	18:44	LB136428
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	18:44	LB136428
	Calcium	234	+/-1000	U	2000	P	07/10/2025	18:44	LB136428
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	18:44	LB136428
	Cobalt	2.26	+/-15	U	30.0	P	07/10/2025	18:44	LB136428
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	18:44	LB136428
	Iron	23.4	+/-50	U	100	P	07/10/2025	18:44	LB136428
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	18:44	LB136428
	Magnesium	244	+/-1000	U	2000	P	07/10/2025	18:44	LB136428
	Manganese	5.94	+/-10	U	20.0	P	07/10/2025	18:44	LB136428
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	18:44	LB136428
	Potassium	918	+/-1000	U	2000	P	07/10/2025	18:44	LB136428
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	18:44	LB136428
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	18:44	LB136428
	Sodium	868	+/-1000	U	2000	P	07/10/2025	18:44	LB136428
	Thallium	4.38	+/-20	U	40.0	P	07/10/2025	18:44	LB136428
	Vanadium	6.26	+/-20	U	40.0	P	07/10/2025	18:44	LB136428
	Zinc	16.7	+/-20	U	40.0	P	07/10/2025	18:44	LB136428
CCB02	Aluminum	11.3	+/-50	U	100	P	07/10/2025	19:34	LB136428
	Antimony	6.76	+/-25	U	50.0	P	07/10/2025	19:34	LB136428
	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	19:34	LB136428
	Barium	14.6	+/-50	U	100	P	07/10/2025	19:34	LB136428
	Beryllium	0.56	+/-3	U	6.00	P	07/10/2025	19:34	LB136428
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	19:34	LB136428
	Calcium	234	+/-1000	U	2000	P	07/10/2025	19:34	LB136428
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	19:34	LB136428
	Cobalt	2.26	+/-15	U	30.0	P	07/10/2025	19:34	LB136428
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	19:34	LB136428
	Iron	23.4	+/-50	U	100	P	07/10/2025	19:34	LB136428
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	19:34	LB136428
	Magnesium	244	+/-1000	U	2000	P	07/10/2025	19:34	LB136428
	Manganese	5.94	+/-10	U	20.0	P	07/10/2025	19:34	LB136428
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	19:34	LB136428
	Potassium	918	+/-1000	U	2000	P	07/10/2025	19:34	LB136428
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	19:34	LB136428

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2543

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	1.62	+/-5	U	10.0	P	07/10/2025	19:34	LB136428
	Sodium	868	+/-1000	U	2000	P	07/10/2025	19:34	LB136428
	Thallium	4.38	+/-20	U	40.0	P	07/10/2025	19:34	LB136428
	Vanadium	6.26	+/-20	U	40.0	P	07/10/2025	19:34	LB136428
	Zinc	16.7	+/-20	U	40.0	P	07/10/2025	19:34	LB136428
CCB03	Aluminum	11.3	+/-50	U	100	P	07/10/2025	20:36	LB136428
	Antimony	6.76	+/-25	U	50.0	P	07/10/2025	20:36	LB136428
	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	20:36	LB136428
	Barium	14.6	+/-50	U	100	P	07/10/2025	20:36	LB136428
	Beryllium	0.56	+/-3	U	6.00	P	07/10/2025	20:36	LB136428
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	20:36	LB136428
	Calcium	234	+/-1000	U	2000	P	07/10/2025	20:36	LB136428
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	20:36	LB136428
	Cobalt	2.26	+/-15	U	30.0	P	07/10/2025	20:36	LB136428
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	20:36	LB136428
	Iron	23.4	+/-50	U	100	P	07/10/2025	20:36	LB136428
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	20:36	LB136428
	Magnesium	244	+/-1000	U	2000	P	07/10/2025	20:36	LB136428
	Manganese	5.94	+/-10	U	20.0	P	07/10/2025	20:36	LB136428
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	20:36	LB136428
	Potassium	918	+/-1000	U	2000	P	07/10/2025	20:36	LB136428
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	20:36	LB136428
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	20:36	LB136428
	Sodium	868	+/-1000	U	2000	P	07/10/2025	20:36	LB136428
	Thallium	4.38	+/-20	U	40.0	P	07/10/2025	20:36	LB136428
	Vanadium	6.26	+/-20	U	40.0	P	07/10/2025	20:36	LB136428
	Zinc	16.7	+/-20	U	40.0	P	07/10/2025	20:36	LB136428
CCB04	Aluminum	11.3	+/-50	U	100	P	07/10/2025	21:31	LB136428
	Antimony	6.76	+/-25	U	50.0	P	07/10/2025	21:31	LB136428
	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	21:31	LB136428
	Barium	14.6	+/-50	U	100	P	07/10/2025	21:31	LB136428
	Beryllium	0.56	+/-3	U	6.00	P	07/10/2025	21:31	LB136428
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	21:31	LB136428
	Calcium	234	+/-1000	U	2000	P	07/10/2025	21:31	LB136428
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	21:31	LB136428
	Cobalt	2.26	+/-15	U	30.0	P	07/10/2025	21:31	LB136428
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	21:31	LB136428
	Iron	23.4	+/-50	U	100	P	07/10/2025	21:31	LB136428
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	21:31	LB136428

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2543

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB04	Magnesium	244	+/-1000	U	2000	P	07/10/2025	21:31	LB136428
	Manganese	5.94	+/-10	U	20.0	P	07/10/2025	21:31	LB136428
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	21:31	LB136428
	Potassium	918	+/-1000	U	2000	P	07/10/2025	21:31	LB136428
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	21:31	LB136428
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	21:31	LB136428
	Sodium	868	+/-1000	U	2000	P	07/10/2025	21:31	LB136428
	Thallium	4.38	+/-20	U	40.0	P	07/10/2025	21:31	LB136428
	Vanadium	6.26	+/-20	U	40.0	P	07/10/2025	21:31	LB136428
	Zinc	16.7	+/-20	U	40.0	P	07/10/2025	21:31	LB136428
CCB05	Aluminum	11.3	+/-50	U	100	P	07/10/2025	22:31	LB136428
	Antimony	6.76	+/-25	U	50.0	P	07/10/2025	22:31	LB136428
	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	22:31	LB136428
	Barium	14.6	+/-50	U	100	P	07/10/2025	22:31	LB136428
	Beryllium	0.56	+/-3	U	6.00	P	07/10/2025	22:31	LB136428
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	22:31	LB136428
	Calcium	234	+/-1000	U	2000	P	07/10/2025	22:31	LB136428
	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	22:31	LB136428
	Cobalt	2.26	+/-15	U	30.0	P	07/10/2025	22:31	LB136428
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	22:31	LB136428
	Iron	23.4	+/-50	U	100	P	07/10/2025	22:31	LB136428
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	22:31	LB136428
	Magnesium	244	+/-1000	U	2000	P	07/10/2025	22:31	LB136428
	Manganese	5.94	+/-10	U	20.0	P	07/10/2025	22:31	LB136428
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	22:31	LB136428
	Potassium	918	+/-1000	U	2000	P	07/10/2025	22:31	LB136428
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	22:31	LB136428
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	22:31	LB136428
	Sodium	868	+/-1000	U	2000	P	07/10/2025	22:31	LB136428
	Thallium	4.38	+/-20	U	40.0	P	07/10/2025	22:31	LB136428
	Vanadium	6.26	+/-20	U	40.0	P	07/10/2025	22:31	LB136428
	Zinc	16.7	+/-20	U	40.0	P	07/10/2025	22:31	LB136428
CCB06	Aluminum	11.3	+/-50	U	100	P	07/10/2025	23:31	LB136428
	Antimony	6.76	+/-25	U	50.0	P	07/10/2025	23:31	LB136428
	Arsenic	5.12	+/-10	U	20.0	P	07/10/2025	23:31	LB136428
	Barium	14.6	+/-50	U	100	P	07/10/2025	23:31	LB136428
	Beryllium	0.56	+/-3	U	6.00	P	07/10/2025	23:31	LB136428
	Cadmium	0.50	+/-3	U	6.00	P	07/10/2025	23:31	LB136428
	Calcium	234	+/-1000	U	2000	P	07/10/2025	23:31	LB136428

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2543

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Chromium	2.12	+/-5	U	10.0	P	07/10/2025	23:31	LB136428
	Cobalt	2.26	+/-15	U	30.0	P	07/10/2025	23:31	LB136428
	Copper	4.60	+/-10	U	20.0	P	07/10/2025	23:31	LB136428
	Iron	23.4	+/-50	U	100	P	07/10/2025	23:31	LB136428
	Lead	2.30	+/-6	U	12.0	P	07/10/2025	23:31	LB136428
	Magnesium	244	+/-1000	U	2000	P	07/10/2025	23:31	LB136428
	Manganese	5.94	+/-10	U	20.0	P	07/10/2025	23:31	LB136428
	Nickel	3.06	+/-20	U	40.0	P	07/10/2025	23:31	LB136428
	Potassium	918	+/-1000	U	2000	P	07/10/2025	23:31	LB136428
	Selenium	9.64	+/-10	U	20.0	P	07/10/2025	23:31	LB136428
	Silver	1.62	+/-5	U	10.0	P	07/10/2025	23:31	LB136428
	Sodium	868	+/-1000	U	2000	P	07/10/2025	23:31	LB136428
	Thallium	4.38	+/-20	U	40.0	P	07/10/2025	23:31	LB136428
	Vanadium	6.26	+/-20	U	40.0	P	07/10/2025	23:31	LB136428
	Zinc	16.7	+/-20	U	40.0	P	07/10/2025	23:31	LB136428
CCB07	Aluminum	11.3	+/-50	U	100	P	07/11/2025	00:22	LB136428
	Antimony	6.76	+/-25	U	50.0	P	07/11/2025	00:22	LB136428
	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	00:22	LB136428
	Barium	14.6	+/-50	U	100	P	07/11/2025	00:22	LB136428
	Beryllium	0.56	+/-3	U	6.00	P	07/11/2025	00:22	LB136428
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	00:22	LB136428
	Calcium	234	+/-1000	U	2000	P	07/11/2025	00:22	LB136428
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	00:22	LB136428
	Cobalt	2.26	+/-15	U	30.0	P	07/11/2025	00:22	LB136428
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	00:22	LB136428
	Iron	23.4	+/-50	U	100	P	07/11/2025	00:22	LB136428
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	00:22	LB136428
	Magnesium	244	+/-1000	U	2000	P	07/11/2025	00:22	LB136428
	Manganese	5.94	+/-10	U	20.0	P	07/11/2025	00:22	LB136428
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	00:22	LB136428
	Potassium	918	+/-1000	U	2000	P	07/11/2025	00:22	LB136428
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	00:22	LB136428
	Silver	1.62	+/-5	U	10.0	P	07/11/2025	00:22	LB136428
	Sodium	868	+/-1000	U	2000	P	07/11/2025	00:22	LB136428
	Thallium	4.38	+/-20	U	40.0	P	07/11/2025	00:22	LB136428
	Vanadium	6.26	+/-20	U	40.0	P	07/11/2025	00:22	LB136428
	Zinc	16.7	+/-20	U	40.0	P	07/11/2025	00:22	LB136428
CCB08	Aluminum	11.3	+/-50	U	100	P	07/11/2025	01:14	LB136428
	Antimony	6.76	+/-25	U	50.0	P	07/11/2025	01:14	LB136428

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2543

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	01:14	LB136428
	Barium	14.6	+/-50	U	100	P	07/11/2025	01:14	LB136428
	Beryllium	0.56	+/-3	U	6.00	P	07/11/2025	01:14	LB136428
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	01:14	LB136428
	Calcium	234	+/-1000	U	2000	P	07/11/2025	01:14	LB136428
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	01:14	LB136428
	Cobalt	2.26	+/-15	U	30.0	P	07/11/2025	01:14	LB136428
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	01:14	LB136428
	Iron	23.4	+/-50	U	100	P	07/11/2025	01:14	LB136428
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	01:14	LB136428
	Magnesium	244	+/-1000	U	2000	P	07/11/2025	01:14	LB136428
	Manganese	5.94	+/-10	U	20.0	P	07/11/2025	01:14	LB136428
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	01:14	LB136428
	Potassium	918	+/-1000	U	2000	P	07/11/2025	01:14	LB136428
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	01:14	LB136428
	Silver	1.62	+/-5	U	10.0	P	07/11/2025	01:14	LB136428
	Sodium	868	+/-1000	U	2000	P	07/11/2025	01:14	LB136428
	Thallium	4.38	+/-20	U	40.0	P	07/11/2025	01:14	LB136428
	Vanadium	6.26	+/-20	U	40.0	P	07/11/2025	01:14	LB136428
	Zinc	16.7	+/-20	U	40.0	P	07/11/2025	01:14	LB136428
CCB09	Aluminum	11.3	+/-50	U	100	P	07/11/2025	02:03	LB136428
	Antimony	6.76	+/-25	U	50.0	P	07/11/2025	02:03	LB136428
	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	02:03	LB136428
	Barium	14.6	+/-50	U	100	P	07/11/2025	02:03	LB136428
	Beryllium	0.56	+/-3	U	6.00	P	07/11/2025	02:03	LB136428
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	02:03	LB136428
	Calcium	234	+/-1000	U	2000	P	07/11/2025	02:03	LB136428
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	02:03	LB136428
	Cobalt	2.26	+/-15	U	30.0	P	07/11/2025	02:03	LB136428
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	02:03	LB136428
	Iron	23.4	+/-50	U	100	P	07/11/2025	02:03	LB136428
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	02:03	LB136428
	Magnesium	244	+/-1000	U	2000	P	07/11/2025	02:03	LB136428
	Manganese	5.94	+/-10	U	20.0	P	07/11/2025	02:03	LB136428
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	02:03	LB136428
	Potassium	918	+/-1000	U	2000	P	07/11/2025	02:03	LB136428
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	02:03	LB136428
	Silver	1.62	+/-5	U	10.0	P	07/11/2025	02:03	LB136428
	Sodium	868	+/-1000	U	2000	P	07/11/2025	02:03	LB136428

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2543

Contract: CAMP02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB09	Thallium	4.38	+/-20	U	40.0	P	07/11/2025	02:03	LB136428
	Vanadium	6.26	+/-20	U	40.0	P	07/11/2025	02:03	LB136428
	Zinc	16.7	+/-20	U	40.0	P	07/11/2025	02:03	LB136428
CCB10	Aluminum	11.3	+/-50	U	100	P	07/11/2025	02:40	LB136428
	Antimony	6.76	+/-25	U	50.0	P	07/11/2025	02:40	LB136428
	Arsenic	5.12	+/-10	U	20.0	P	07/11/2025	02:40	LB136428
	Barium	14.6	+/-50	U	100	P	07/11/2025	02:40	LB136428
	Beryllium	0.56	+/-3	U	6.00	P	07/11/2025	02:40	LB136428
	Cadmium	0.50	+/-3	U	6.00	P	07/11/2025	02:40	LB136428
	Calcium	234	+/-1000	U	2000	P	07/11/2025	02:40	LB136428
	Chromium	2.12	+/-5	U	10.0	P	07/11/2025	02:40	LB136428
	Cobalt	2.26	+/-15	U	30.0	P	07/11/2025	02:40	LB136428
	Copper	4.60	+/-10	U	20.0	P	07/11/2025	02:40	LB136428
	Iron	23.4	+/-50	U	100	P	07/11/2025	02:40	LB136428
	Lead	2.30	+/-6	U	12.0	P	07/11/2025	02:40	LB136428
	Magnesium	244	+/-1000	U	2000	P	07/11/2025	02:40	LB136428
	Manganese	5.94	+/-10	U	20.0	P	07/11/2025	02:40	LB136428
	Nickel	3.06	+/-20	U	40.0	P	07/11/2025	02:40	LB136428
	Potassium	918	+/-1000	U	2000	P	07/11/2025	02:40	LB136428
	Selenium	9.64	+/-10	U	20.0	P	07/11/2025	02:40	LB136428
	Silver	1.62	+/-5	U	10.0	P	07/11/2025	02:40	LB136428
	Sodium	868	+/-1000	U	2000	P	07/11/2025	02:40	LB136428
	Thallium	4.38	+/-20	U	40.0	P	07/11/2025	02:40	LB136428
	Vanadium	6.26	+/-20	U	40.0	P	07/11/2025	02:40	LB136428
	Zinc	16.7	+/-20	U	40.0	P	07/11/2025	02:40	LB136428