

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2543

Instrument: CV1

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB168821BL		SOLID		Batch Number:	PB168821		Prep Date:	07/10/2025	
	Mercury	0.0080	<0.013	U	0.013	CV	07/11/2025	14:35	LB136438

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: CDM Smith

SDG No.: Q2543

Instrument: P4

Sample ID	Analyte	Result (mg/Kg)	Acceptance Limit	Conc Qual	CRQL mg/Kg	M	Analysis Date	Analysis Time	Run
PB168790BL	SOLID			Batch Number:	PB168790		Prep Date:	07/10/2025	
	Aluminum	0.84	<2.5	U	5.00	P	07/11/2025	01:18	LB136428
	Antimony	0.22	<1.25	U	2.50	P	07/11/2025	01:18	LB136428
	Arsenic	0.19	<0.5	U	1.00	P	07/11/2025	01:18	LB136428
	Barium	0.73	<2.5	U	5.00	P	07/11/2025	01:18	LB136428
	Beryllium	0.025	<0.15	U	0.30	P	07/11/2025	01:18	LB136428
	Cadmium	0.024	<0.15	U	0.30	P	07/11/2025	01:18	LB136428
	Calcium	11.1	<50	U	100	P	07/11/2025	01:18	LB136428
	Chromium	0.047	<0.25	U	0.50	P	07/11/2025	01:18	LB136428
	Cobalt	0.10	<0.75	U	1.50	P	07/11/2025	01:18	LB136428
	Copper	0.22	<0.5	U	1.00	P	07/11/2025	01:18	LB136428
	Iron	3.99	<2.5	U	5.00	P	07/11/2025	01:18	LB136428
	Lead	0.13	<0.3	U	0.60	P	07/11/2025	01:18	LB136428
	Magnesium	12.0	<50	U	100	P	07/11/2025	01:18	LB136428
	Manganese	0.14	<0.5	U	1.00	P	07/11/2025	01:18	LB136428
	Nickel	0.13	<1	U	2.00	P	07/11/2025	01:18	LB136428
	Potassium	27.7	<50	U	100	P	07/11/2025	01:18	LB136428
	Selenium	0.26	<0.5	U	1.00	P	07/11/2025	01:18	LB136428
	Silver	0.12	<0.25	U	0.50	P	07/11/2025	01:18	LB136428
	Sodium	17.8	<50	U	100	P	07/11/2025	01:18	LB136428
	Thallium	0.23	<1	U	2.00	P	07/11/2025	01:18	LB136428
	Vanadium	0.25	<1	U	2.00	P	07/11/2025	01:18	LB136428
	Zinc	0.23	<1	U	2.00	P	07/11/2025	01:18	LB136428