

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB062325

Review By	yogesh	Review On	6/23/2025 3:07:11 PM
Supervise By	mohammad	Supervise On	6/25/2025 2:41:46 AM
SubDirectory	FB062325	HP Acquire Method	HP Processing Method
			FB062325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24656,PP24658,PP24659,PP24660,PP24661		
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24662		

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	5 GRO STD		FB031909.D	23 Jun 2025 12:17		YP/AJ	Ok
2	10 GRO STD		FB031910.D	23 Jun 2025 12:44		YP/AJ	Ok
3	20 GRO STD		FB031911.D	23 Jun 2025 13:12		YP/AJ	Ok
4	50 GRO STD		FB031912.D	23 Jun 2025 13:40		YP/AJ	Ok
5	100 GRO STD		FB031913.D	23 Jun 2025 14:07		YP/AJ	Ok
6	FB062325GROICV		FB031914.D	23 Jun 2025 14:48		YP/AJ	Ok

M : Manual Integration

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB071025

Review By	yogesh	Review On	7/10/2025 3:17:40 PM
Supervise By	mohammad	Supervise On	7/11/2025 9:10:17 AM
SubDirectory	FB071025	HP Acquire Method	HP Processing Method
			FB062325

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	PP24110,PP24656,PP24658,PP24659,PP24660,PP24661
CCC Internal Standard/PEM	PP24710,PP24711,PP24712
ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	PP24111,PP24662

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	20 PPB GRO STD		FB032059.D	10 Jul 2025 9:08		YP/AJ	Ok,M
2	VBF0710S1		FB032060.D	10 Jul 2025 9:50		YP/AJ	Ok
3	VBF0710S2		FB032061.D	10 Jul 2025 10:18		YP/AJ	Ok
4	BSF0710S1		FB032062.D	10 Jul 2025 10:45		YP/AJ	Ok
5	Q2487-15		FB032063.D	10 Jul 2025 11:25	Vial-B, Not Purged	YP/AJ	Not Ok
6	Q2487-16		FB032064.D	10 Jul 2025 11:53	Vial-B, Not Purged	YP/AJ	Not Ok
7	Q2543-01		FB032065.D	10 Jul 2025 12:36	Vial-A	YP/AJ	Ok
8	Q2543-02		FB032066.D	10 Jul 2025 13:04	Vial-A, Surr Fail	YP/AJ	Not Ok
9	Q2543-03		FB032067.D	10 Jul 2025 13:31	Vial-A	YP/AJ	Ok
10	Q2543-04		FB032068.D	10 Jul 2025 13:59	Vial-A	YP/AJ	Ok
11	Q2487-15		FB032069.D	10 Jul 2025 14:27	Vial-C	YP/AJ	Ok
12	BSF0710S2		FB032070.D	10 Jul 2025 14:54		YP/AJ	Ok
13	20 PPB GRO STD		FB032071.D	10 Jul 2025 15:22		YP/AJ	Ok
14	Q2487-16		FB032072.D	10 Jul 2025 16:24	Vial-C	YP/AJ	Ok
15	Q2543-02		FB032073.D	10 Jul 2025 16:52	Vial-B	YP/AJ	Ok
16	Q2529-01		FB032074.D	10 Jul 2025 17:20	Vial-A	YP/AJ	Ok
17	Q2529-02		FB032075.D	10 Jul 2025 17:48	Vial-A	YP/AJ	Ok
18	Q2529-03		FB032076.D	10 Jul 2025 18:15	Vial-A, Not Purged	YP/AJ	Not Ok

Instrument ID: FID_B

Daily Analysis Runlog For Sequence/QC Batch ID # FB071025

Review By	yogesh	Review On	7/10/2025 3:17:40 PM
Supervise By	mohammad	Supervise On	7/11/2025 9:10:17 AM
SubDirectory	FB071025	HP Acquire Method	HP Processing Method
			FB062325
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24110,PP24656,PP24658,PP24659,PP24660,PP24661		
CCC	PP24710,PP24711,PP24712		
Internal Standard/PEM			
ICV/I.BLK	PP24111,PP24662		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	Q2529-04		FB032077.D	10 Jul 2025 18:43	Vial-A	YP/AJ	Ok
20	Q2529-05		FB032078.D	10 Jul 2025 19:10	Vial-A	YP/AJ	Ok
21	Q2529-06		FB032079.D	10 Jul 2025 19:38	Vial-A, Not Purged	YP/AJ	Not Ok
22	Q2529-07		FB032080.D	10 Jul 2025 20:05	Vial-A	YP/AJ	Ok
23	Q2529-08		FB032081.D	10 Jul 2025 20:33	Vial-A, Not Purged	YP/AJ	Not Ok
24	Q2529-09		FB032082.D	10 Jul 2025 21:00	Vial-A, Not Purged	YP/AJ	Not Ok
25	Q2529-10		FB032083.D	10 Jul 2025 21:28	Vial-A	YP/AJ	Ok
26	BSF0710S3		FB032084.D	10 Jul 2025 21:55		YP/AJ	Ok,M
27	20 PPB GRO STD		FB032085.D	10 Jul 2025 23:17		YP/AJ	Ok,M

M : Manual Integration