

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2543</u>	Analytical Method:	<u>8151A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PS031145.D</u>

	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	
			Result	Result			Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2529-10MS (Column 1)	Client Sample ID:		TP-30MS								
	DICAMBA	182.3	0	83.6	ug/Kg	46				10	112	
	DICHLORPROP	182.3	0	84.5	ug/Kg	46				10	113	
	2,4-D	182.3	0	179	ug/Kg	98				10	144	
	2,4,5-TP(Silvex)	182.3	0	98.3	ug/Kg	54				10	114	
	2,4,5-T	182.3	0	121	ug/Kg	66				10	115	
	2,4-DB	182.3	0	30.0	ug/Kg	16				10	140	
	Dinoseb	182.3	0	0	ug/Kg	0	*			10	118	
Lab Sample ID:	Q2529-10MS (Column 2)	Client Sample ID:		TP-30MS								
	DICAMBA	182.3	0	82.7	ug/Kg	45				10	112	
	DICHLORPROP	182.3	0	90.6	ug/Kg	50				10	113	
	2,4-D	182.3	0	152	ug/Kg	83				10	144	
	2,4,5-TP(Silvex)	182.3	0	284	ug/Kg	156				10	114	
	2,4,5-T	182.3	0	85.0	ug/Kg	47				10	115	
	2,4-DB	182.3	0	29.1	ug/Kg	16				10	140	
	Dinoseb	182.3	0	0	ug/Kg	0	*			10	118	