

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.:	<u>Q2543</u>	Analytical Method:	<u>8151A</u>
Client:	<u>CDM Smith</u>	DataFile :	<u>PS031146.D</u>

		Sample				Rec		RPD		Limits		
	Parameter	Spike	Result	Result	Units	Rec	Qual	RPD	Qual	Low	High	RPD
Lab Sample ID:	Q2529-10MSD (Column 1)	Client Sample ID:		TP-30MSD								
	DICAMBA	182	0	98.6	ug/Kg	54		16		10	112	20
	DICHLORPROP	182	0	102	ug/Kg	56		20		10	113	20
	2,4-D	182	0	192	ug/Kg	105		7		10	144	20
	2,4,5-TP(Silvex)	182	0	117	ug/Kg	64		17		10	114	20
	2,4,5-T	182	0	143	ug/Kg	79		18		10	115	20
	2,4-DB	182	0	32.0	ug/Kg	18		12		10	140	20
	Dinoseb	182	0	0	ug/Kg	0	*	0		10	118	20
Lab Sample ID:	Q2529-10MSD (Column 2)	Client Sample ID:		TP-30MSD								
	DICAMBA	182	0	98.1	ug/Kg	54		18		10	112	20
	DICHLORPROP	182	0	107	ug/Kg	59		17		10	113	20
	2,4-D	182	0	164	ug/Kg	90		8		10	144	20
	2,4,5-TP(Silvex)	182	0	297	ug/Kg	163		4		10	114	20
	2,4,5-T	182	0	101	ug/Kg	55		16		10	115	20
	2,4-DB	182	0	29.0	ug/Kg	16		0		10	140	20
	Dinoseb	182	0	0	ug/Kg	0	*	0		10	118	20