

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2543
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112181.D **Time Analyzed:** 09:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.758	4.661	4.861	499.430	500.000	-0.1
Aroclor-1016-2	4.776	4.680	4.880	513.210	500.000	2.6
Aroclor-1016-3	4.833	4.736	4.936	498.860	500.000	-0.2
Aroclor-1016-4	4.953	4.856	5.056	510.770	500.000	2.2
Aroclor-1016-5	5.210	5.113	5.313	515.740	500.000	3.1
Aroclor-1260-1	6.248	6.151	6.351	497.790	500.000	-0.4
Aroclor-1260-2	6.437	6.341	6.541	503.840	500.000	0.8
Aroclor-1260-3	6.803	6.707	6.907	498.140	500.000	-0.4
Aroclor-1260-4	7.063	6.967	7.167	519.110	500.000	3.8
Aroclor-1260-5	7.306	7.209	7.409	534.810	500.000	7.0
Decachlorobiphenyl	8.699	8.603	8.803	53.080	50.000	6.2
Tetrachloro-m-xylene	3.670	3.573	3.773	52.160	50.000	4.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2543
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO112181.D **Time Analyzed:** 09:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.740	4.661	4.861	459.390	500.000	-8.1
Aroclor-1016-2	4.759	4.679	4.879	466.230	500.000	-6.8
Aroclor-1016-3	4.934	4.854	5.054	476.460	500.000	-4.7
Aroclor-1016-4	4.977	4.897	5.097	472.030	500.000	-5.6
Aroclor-1016-5	5.189	5.109	5.309	497.890	500.000	-0.4
Aroclor-1260-1	6.218	6.138	6.338	489.330	500.000	-2.1
Aroclor-1260-2	6.406	6.326	6.526	492.960	500.000	-1.4
Aroclor-1260-3	6.558	6.478	6.678	505.180	500.000	1.0
Aroclor-1260-4	7.028	6.948	7.148	472.680	500.000	-5.5
Aroclor-1260-5	7.269	7.190	7.390	474.810	500.000	-5.0
Decachlorobiphenyl	8.647	8.568	8.768	50.630	50.000	1.3
Tetrachloro-m-xylene	3.665	3.584	3.784	49.470	50.000	-1.1