

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CAMP02  
**Lab Code:** ACE **SDG NO.:** Q2543  
**GC Column:** ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 07/10/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO112196.D **Time Analyzed:** 17:18

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.759 | 4.661     | 4.861 | 560.040            | 500.000           | 12.0  |
| Aroclor-1016-2       | 4.777 | 4.680     | 4.880 | 560.420            | 500.000           | 12.1  |
| Aroclor-1016-3       | 4.834 | 4.736     | 4.936 | 543.320            | 500.000           | 8.7   |
| Aroclor-1016-4       | 4.954 | 4.856     | 5.056 | 558.410            | 500.000           | 11.7  |
| Aroclor-1016-5       | 5.212 | 5.113     | 5.313 | 536.860            | 500.000           | 7.4   |
| Aroclor-1260-1       | 6.249 | 6.151     | 6.351 | 475.630            | 500.000           | -4.9  |
| Aroclor-1260-2       | 6.439 | 6.341     | 6.541 | 413.920            | 500.000           | -17.2 |
| Aroclor-1260-3       | 6.805 | 6.707     | 6.907 | 406.480            | 500.000           | -18.7 |
| Aroclor-1260-4       | 7.064 | 6.967     | 7.167 | 472.020            | 500.000           | -5.6  |
| Aroclor-1260-5       | 7.308 | 7.209     | 7.409 | 430.670            | 500.000           | -13.9 |
| Decachlorobiphenyl   | 8.701 | 8.603     | 8.803 | 38.770             | 50.000            | -22.5 |
| Tetrachloro-m-xylene | 3.672 | 3.573     | 3.773 | 59.100             | 50.000            | 18.2  |

### CALIBRATION VERIFICATION SUMMARY

**Lab Name:** Alliance **Contract:** CAMP02  
**Lab Code:** ACE **SDG NO.:** Q2543  
**GC Column:** ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/08/2025 07/08/2025

**Client Sample No.:** CCAL02 **Date Analyzed:** 07/10/2025  
**Lab Sample No.:** AR1660CCC500 **Data File :** PO112196.D **Time Analyzed:** 17:18

| COMPOUND             | RT    | RT WINDOW |       | CALC<br>AMOUNT(ng) | NOM<br>AMOUNT(ng) | %D    |
|----------------------|-------|-----------|-------|--------------------|-------------------|-------|
|                      |       | FROM      | TO    |                    |                   |       |
| Aroclor-1016-1       | 4.740 | 4.661     | 4.861 | 537.020            | 500.000           | 7.4   |
| Aroclor-1016-2       | 4.759 | 4.679     | 4.879 | 515.100            | 500.000           | 3.0   |
| Aroclor-1016-3       | 4.934 | 4.854     | 5.054 | 531.360            | 500.000           | 6.3   |
| Aroclor-1016-4       | 4.977 | 4.897     | 5.097 | 515.570            | 500.000           | 3.1   |
| Aroclor-1016-5       | 5.188 | 5.109     | 5.309 | 554.910            | 500.000           | 11.0  |
| Aroclor-1260-1       | 6.218 | 6.138     | 6.338 | 462.920            | 500.000           | -7.4  |
| Aroclor-1260-2       | 6.406 | 6.326     | 6.526 | 459.680            | 500.000           | -8.1  |
| Aroclor-1260-3       | 6.558 | 6.478     | 6.678 | 452.360            | 500.000           | -9.5  |
| Aroclor-1260-4       | 7.027 | 6.948     | 7.148 | 420.160            | 500.000           | -16.0 |
| Aroclor-1260-5       | 7.269 | 7.190     | 7.390 | 432.610            | 500.000           | -13.5 |
| Decachlorobiphenyl   | 8.648 | 8.568     | 8.768 | 44.850             | 50.000            | -10.3 |
| Tetrachloro-m-xylene | 3.664 | 3.584     | 3.784 | 56.840             | 50.000            | 13.7  |