

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2543
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073691.D **Time Analyzed:** 09:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.638	5.539	5.739	472.060	500.000	-5.6
Aroclor-1016-2	5.659	5.560	5.760	493.380	500.000	-1.3
Aroclor-1016-3	5.721	5.622	5.822	489.480	500.000	-2.1
Aroclor-1016-4	5.819	5.720	5.920	499.970	500.000	0.0
Aroclor-1016-5	6.111	6.012	6.212	522.700	500.000	4.5
Aroclor-1260-1	7.228	7.129	7.329	491.200	500.000	-1.8
Aroclor-1260-2	7.481	7.383	7.583	421.210	500.000	-15.8
Aroclor-1260-3	7.839	7.740	7.940	468.730	500.000	-6.3
Aroclor-1260-4	8.063	7.964	8.164	484.890	500.000	-3.0
Aroclor-1260-5	8.380	8.283	8.483	477.770	500.000	-4.4
Decachlorobiphenyl	10.174	10.076	10.276	50.150	50.000	0.3
Tetrachloro-m-xylene	4.488	4.388	4.588	50.060	50.000	0.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2543
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL03 **Date Analyzed:** 07/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073691.D **Time Analyzed:** 09:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.856	4.758	4.958	502.710	500.000	0.5
Aroclor-1016-2	4.873	4.775	4.975	500.290	500.000	0.1
Aroclor-1016-3	5.050	4.951	5.151	509.040	500.000	1.8
Aroclor-1016-4	5.091	4.993	5.193	505.520	500.000	1.1
Aroclor-1016-5	5.305	5.207	5.407	528.100	500.000	5.6
Aroclor-1260-1	6.335	6.237	6.437	522.390	500.000	4.5
Aroclor-1260-2	6.524	6.425	6.625	534.750	500.000	7.0
Aroclor-1260-3	6.675	6.577	6.777	521.300	500.000	4.3
Aroclor-1260-4	7.144	7.046	7.246	532.670	500.000	6.5
Aroclor-1260-5	7.386	7.289	7.489	525.960	500.000	5.2
Decachlorobiphenyl	8.777	8.680	8.880	57.890	50.000	15.8
Tetrachloro-m-xylene	3.778	3.678	3.878	50.320	50.000	0.6