

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2543
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073705.D **Time Analyzed:** 15:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.638	5.539	5.739	457.300	500.000	-8.5
Aroclor-1016-2	5.660	5.560	5.760	473.030	500.000	-5.4
Aroclor-1016-3	5.722	5.622	5.822	477.830	500.000	-4.4
Aroclor-1016-4	5.819	5.720	5.920	490.940	500.000	-1.8
Aroclor-1016-5	6.111	6.012	6.212	498.030	500.000	-0.4
Aroclor-1260-1	7.228	7.129	7.329	457.580	500.000	-8.5
Aroclor-1260-2	7.482	7.383	7.583	396.380	500.000	-20.7
Aroclor-1260-3	7.840	7.740	7.940	442.240	500.000	-11.6
Aroclor-1260-4	8.063	7.964	8.164	447.850	500.000	-10.4
Aroclor-1260-5	8.381	8.283	8.483	458.600	500.000	-8.3
Decachlorobiphenyl	10.174	10.076	10.276	47.000	50.000	-6.0
Tetrachloro-m-xylene	4.488	4.388	4.588	50.640	50.000	1.3

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2543
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 07/07/2025 07/07/2025

Client Sample No.: CCAL04 **Date Analyzed:** 07/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP073705.D **Time Analyzed:** 15:50

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.856	4.758	4.958	511.240	500.000	2.2
Aroclor-1016-2	4.874	4.775	4.975	511.870	500.000	2.4
Aroclor-1016-3	5.051	4.951	5.151	514.240	500.000	2.8
Aroclor-1016-4	5.092	4.993	5.193	508.640	500.000	1.7
Aroclor-1016-5	5.306	5.207	5.407	555.300	500.000	11.1
Aroclor-1260-1	6.335	6.237	6.437	473.360	500.000	-5.3
Aroclor-1260-2	6.524	6.425	6.625	481.990	500.000	-3.6
Aroclor-1260-3	6.676	6.577	6.777	455.070	500.000	-9.0
Aroclor-1260-4	7.145	7.046	7.246	462.000	500.000	-7.6
Aroclor-1260-5	7.387	7.289	7.489	457.020	500.000	-8.6
Decachlorobiphenyl	8.778	8.680	8.880	50.480	50.000	1.0
Tetrachloro-m-xylene	3.778	3.678	3.878	51.390	50.000	2.8