



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q2543</u>	Analytical Method:	<u>8082A</u>
Client:	<u>CDM Smith</u>	Datafile :	<u>PP073701.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB168783BS (Column 1)	AR1016	166.6	152	ug/kg	91				71	120	
	AR1260	166.6	142	ug/kg	85				65	130	
PB168783BS (Column 2)	AR1016	166.6	157	ug/kg	94				71	120	
	AR1260	166.6	157	ug/kg	94				65	130	