

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2543
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089432.D **Time Analyzed:** 12:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	51.350	50.000	2.7
4,4'-DDE	6.195	6.096	6.296	50.690	50.000	1.4
4,4'-DDT	7.019	6.921	7.121	49.590	50.000	-0.8
Aldrin	5.270	5.171	5.371	51.560	50.000	3.1
alpha-BHC	3.998	3.899	4.099	52.080	50.000	4.2
alpha-Chlordane	6.026	5.927	6.127	50.700	50.000	1.4
beta-BHC	4.515	4.415	4.615	50.030	50.000	0.1
Decachlorobiphenyl	9.069	8.972	9.172	49.450	50.000	-1.1
delta-BHC	4.763	4.664	4.864	54.450	50.000	8.9
Dieldrin	6.346	6.247	6.447	51.040	50.000	2.1
Endosulfan I	6.073	5.975	6.175	50.780	50.000	1.6
Endosulfan II	6.785	6.686	6.886	51.070	50.000	2.1
Endosulfan sulfate	7.148	7.049	7.249	49.480	50.000	-1.0
Endrin	6.573	6.475	6.675	48.550	50.000	-2.9
Endrin aldehyde	6.914	6.815	7.015	48.380	50.000	-3.2
Endrin ketone	7.628	7.530	7.730	50.230	50.000	0.5
gamma-BHC (Lindane)	4.329	4.230	4.430	51.710	50.000	3.4
gamma-Chlordane	5.945	5.846	6.046	51.120	50.000	2.2
Heptachlor	4.928	4.829	5.029	51.590	50.000	3.2
Heptachlor epoxide	5.689	5.591	5.791	50.170	50.000	0.3
Methoxychlor	7.492	7.393	7.593	46.600	50.000	-6.8
Tetrachloro-m-xylene	3.549	3.450	3.650	51.680	50.000	3.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2543
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL01 **Date Analyzed:** 07/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089432.D **Time Analyzed:** 12:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.927	5.830	6.030	52.780	50.000	5.6
4,4'-DDE	5.373	5.275	5.475	53.420	50.000	6.8
4,4'-DDT	6.181	6.084	6.284	51.610	50.000	3.2
Aldrin	4.366	4.269	4.469	53.750	50.000	7.5
alpha-BHC	3.391	3.293	3.493	53.650	50.000	7.3
alpha-Chlordane	5.188	5.091	5.291	53.190	50.000	6.4
beta-BHC	4.023	3.926	4.126	52.780	50.000	5.6
Decachlorobiphenyl	8.070	7.972	8.172	53.860	50.000	7.7
delta-BHC	4.260	4.162	4.362	53.590	50.000	7.2
Dieldrin	5.510	5.413	5.613	53.060	50.000	6.1
Endosulfan I	5.244	5.147	5.347	53.770	50.000	7.5
Endosulfan II	6.078	5.981	6.181	52.450	50.000	4.9
Endosulfan sulfate	6.479	6.383	6.583	52.420	50.000	4.8
Endrin	5.787	5.689	5.889	50.680	50.000	1.4
Endrin aldehyde	6.256	6.159	6.359	50.520	50.000	1.0
Endrin ketone	6.989	6.892	7.092	52.660	50.000	5.3
gamma-BHC (Lindane)	3.726	3.630	3.830	53.590	50.000	7.2
gamma-Chlordane	5.122	5.026	5.226	52.900	50.000	5.8
Heptachlor	4.081	3.983	4.183	52.390	50.000	4.8
Heptachlor epoxide	4.870	4.773	4.973	55.240	50.000	10.5
Methoxychlor	6.752	6.655	6.855	49.380	50.000	-1.2
Tetrachloro-m-xylene	2.879	2.780	2.980	54.960	50.000	9.9