

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2543
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089448.D **Time Analyzed:** 19:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.704	6.605	6.805	48.150	50.000	-3.7
4,4'-DDE	6.194	6.096	6.296	50.330	50.000	0.7
4,4'-DDT	7.020	6.921	7.121	43.790	50.000	-12.4
Aldrin	5.270	5.171	5.371	53.130	50.000	6.3
alpha-BHC	3.999	3.899	4.099	53.870	50.000	7.7
alpha-Chlordane	6.026	5.927	6.127	50.530	50.000	1.1
beta-BHC	4.515	4.415	4.615	51.900	50.000	3.8
Decachlorobiphenyl	9.071	8.972	9.172	49.130	50.000	-1.7
delta-BHC	4.763	4.664	4.864	56.310	50.000	12.6
Dieldrin	6.346	6.247	6.447	50.360	50.000	0.7
Endosulfan I	6.073	5.975	6.175	50.400	50.000	0.8
Endosulfan II	6.785	6.686	6.886	46.770	50.000	-6.5
Endosulfan sulfate	7.149	7.049	7.249	46.270	50.000	-7.5
Endrin	6.573	6.475	6.675	45.900	50.000	-8.2
Endrin aldehyde	6.914	6.815	7.015	45.350	50.000	-9.3
Endrin ketone	7.629	7.530	7.730	44.450	50.000	-11.1
gamma-BHC (Lindane)	4.330	4.230	4.430	53.480	50.000	7.0
gamma-Chlordane	5.945	5.846	6.046	51.250	50.000	2.5
Heptachlor	4.929	4.829	5.029	51.830	50.000	3.7
Heptachlor epoxide	5.689	5.591	5.791	50.600	50.000	1.2
Methoxychlor	7.492	7.393	7.593	40.060	50.000	-19.9
Tetrachloro-m-xylene	3.550	3.450	3.650	51.250	50.000	2.5

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CAMP02
Lab Code: ACE **SDG NO.:** Q2543
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 06/17/2025 06/17/2025

Client Sample No.: CCAL02 **Date Analyzed:** 07/10/2025
Lab Sample No.: PSTDCCC050 **Data File :** PD089448.D **Time Analyzed:** 19:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.928	5.830	6.030	52.630	50.000	5.3
4,4'-DDE	5.373	5.275	5.475	53.720	50.000	7.4
4,4'-DDT	6.181	6.084	6.284	47.410	50.000	-5.2
Aldrin	4.367	4.269	4.469	55.150	50.000	10.3
alpha-BHC	3.393	3.293	3.493	55.660	50.000	11.3
alpha-Chlordane	5.189	5.091	5.291	53.940	50.000	7.9
beta-BHC	4.024	3.926	4.126	54.460	50.000	8.9
Decachlorobiphenyl	8.070	7.972	8.172	44.480	50.000	-11.0
delta-BHC	4.261	4.162	4.362	55.350	50.000	10.7
Dieldrin	5.511	5.413	5.613	53.550	50.000	7.1
Endosulfan I	5.246	5.147	5.347	54.160	50.000	8.3
Endosulfan II	6.078	5.981	6.181	50.800	50.000	1.6
Endosulfan sulfate	6.480	6.383	6.583	49.420	50.000	-1.2
Endrin	5.787	5.689	5.889	51.950	50.000	3.9
Endrin aldehyde	6.257	6.159	6.359	48.410	50.000	-3.2
Endrin ketone	6.990	6.892	7.092	46.330	50.000	-7.3
gamma-BHC (Lindane)	3.727	3.630	3.830	55.360	50.000	10.7
gamma-Chlordane	5.124	5.026	5.226	54.290	50.000	8.6
Heptachlor	4.082	3.983	4.183	53.040	50.000	6.1
Heptachlor epoxide	4.871	4.773	4.973	55.620	50.000	11.2
Methoxychlor	6.752	6.655	6.855	43.270	50.000	-13.5
Tetrachloro-m-xylene	2.881	2.780	2.980	54.620	50.000	9.2