

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: CHEM

SDG NO.: Q2543

Client ID : SSTDICCC040

Date Analyzed: 07/08/2025

Lab File ID: BM050381.D

Time Analyzed: 15:20

Instrument ID: BNA_M

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	390087	7.863	1498320	10.66	967077	14.49
UPPER LIMIT	780174	8.363	2996640	11.163	1934150	14.992
LOWER LIMIT	195044	7.363	749160	10.163	483539	13.992
EPA SAMPLE NO.						
01 TP-49MS	627832	7.86	2379650	10.66	1542630	14.49
02 TP-49MSD	614645	7.86	2249890	10.66	1411020	14.49
03 TP-23	581438	7.86	2127800	10.66	1370950	14.49
04 TP-23-99	639438	7.86	2330150	10.66	1499850	14.49
05 TP-49	577892	7.86	2139760	10.66	1403470	14.49

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.