

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2543

Client ID : SSTDCCC040

Date Analyzed: 07/11/2025

Lab File ID: BP025098.D

Time Analyzed: 10:26

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	436580	7.437	1693300	10.18	1072860	14.10
UPPER LIMIT	873160	7.937	3386600	10.684	2145720	14.595
LOWER LIMIT	218290	6.937	846650	9.684	536430	13.595
EPA SAMPLE NO.						
01 PB168787BL	435629	7.44	1705700	10.18	1062770	14.08

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.