

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q2543

Client ID : SSTDCCC040

Date Analyzed: 07/15/2025

Lab File ID: BP025138.D

Time Analyzed: 09:36

Instrument ID: BNA_P

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	539286	7.443	1912690	10.18	1092260	14.08
UPPER LIMIT	1078570	7.943	3825380	10.684	2184520	14.584
LOWER LIMIT	269643	6.943	956345	9.684	546130	13.584
EPA SAMPLE NO.						
01 PB168787BS	578947	7.44	2205740	10.18	1344570	14.08
02 TP-41	556317	7.44	2503670	10.19	1662470	14.10

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.