

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM070925

Review By	Rahul	Review On	7/9/2025 9:48:51 AM
Supervise By	Jagrut	Supervise On	7/9/2025 2:15:39 PM
SubDirectory	BM070925	HP Acquire Method	BNA_M
		HP Processing Method	bm070925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050376.D	08 Jul 2025 11:59		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BM050377.D	08 Jul 2025 12:39		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BM050378.D	08 Jul 2025 13:19		RC/JU	Ok
4	SSTDICC010	SSTDICC010	BM050379.D	08 Jul 2025 14:00		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BM050380.D	08 Jul 2025 14:40		RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BM050381.D	08 Jul 2025 15:20	Compound#32,54,57,62,65 Kept on LR	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BM050382.D	08 Jul 2025 16:01		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BM050383.D	08 Jul 2025 16:41		RC/JU	Ok
9	SSTDICC080	SSTDICC080	BM050384.D	08 Jul 2025 17:22	Compound#09 removed from 80 ppm	RC/JU	Ok
10	SSTDICV040	ICVBM070925	BM050385.D	08 Jul 2025 18:05	Comp#77 Failed from High side	RC/JU	Ok,M
11	PB168722BL	PB168722BL	BM050386.D	08 Jul 2025 19:26		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_M

Daily Analysis Runlog For Sequence/QC Batch ID # BM071025

Review By	Rahul	Review On	7/11/2025 12:08:59 PM
Supervise By	Jagrut	Supervise On	7/11/2025 2:56:38 PM
SubDirectory	BM071025	HP Acquire Method	BNA_M
		HP Processing Method	bm070925

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BM050387.D	10 Jul 2025 13:13		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BM050388.D	10 Jul 2025 13:53		RC/JU	Ok,M
3	PB168689TB	PB168689TB	BM050389.D	10 Jul 2025 14:33		RC/JU	Ok
4	Q2529-03	TP-79	BM050390.D	10 Jul 2025 15:17		RC/JU	Ok
5	Q2529-05	TP-98	BM050391.D	10 Jul 2025 15:57		RC/JU	Ok
6	Q2529-08	TP-89	BM050392.D	10 Jul 2025 16:37		RC/JU	Ok
7	Q2529-09	TP-33	BM050393.D	10 Jul 2025 17:18		RC/JU	Ok
8	Q2543-02	TP-49	BM050394.D	10 Jul 2025 17:58		RC/JU	Ok,M
9	Q2543-02MS	TP-49MS	BM050395.D	10 Jul 2025 18:38		RC/JU	Ok,M
10	Q2543-02MSD	TP-49MSD	BM050396.D	10 Jul 2025 19:18		RC/JU	Ok,M
11	Q2489-01	G4(0-6)	BM050397.D	10 Jul 2025 19:59		RC/JU	Ok
12	Q2489-01MS	G4(0-6)MS	BM050398.D	10 Jul 2025 20:39		RC/JU	Ok,M
13	Q2489-01MSD	G4(0-6)MSD	BM050399.D	10 Jul 2025 21:19		RC/JU	Ok,M
14	Q2489-02	G4(6-12)	BM050400.D	10 Jul 2025 21:59		RC/JU	Ok
15	Q2489-03	G3(0-6)	BM050401.D	10 Jul 2025 22:39		RC/JU	Ok
16	Q2543-03	TP-23	BM050402.D	10 Jul 2025 23:19		RC/JU	Ok
17	Q2543-04	TP-23-99	BM050403.D	10 Jul 2025 23:59		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP070325

Review By	Rahul	Review On	7/7/2025 11:36:11 AM
Supervise By	Jagrut	Supervise On	7/7/2025 12:15:03 PM
SubDirectory	BP070325	HP Acquire Method	BNA_P
		HP Processing Method	bp070325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12672,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025070.D	03 Jul 2025 08:58		CG/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP025071.D	03 Jul 2025 09:39		CG/JU	Ok
3	SSTDICC005	SSTDICC005	BP025072.D	03 Jul 2025 10:21		CG/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP025073.D	03 Jul 2025 11:02		CG/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP025074.D	03 Jul 2025 11:44		CG/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP025075.D	03 Jul 2025 12:25	compound #77 failed in the calibration.	CG/JU	Ok,M
7	SSTDICC050	SSTDICC050	BP025076.D	03 Jul 2025 13:07	Compound#26,32,48,54,57,65,80,85 Kept on LR	CG/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP025077.D	03 Jul 2025 13:49		CG/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP025078.D	03 Jul 2025 14:31	Compound#09 removed from 80 ppm	CG/JU	Ok,M
10	SSTDICV040	ICVBP070325	BP025079.D	03 Jul 2025 16:18		CG/JU	Ok,M
11	PB168694BL	PB168694BL	BP025080.D	03 Jul 2025 16:59	Analyzed for contamination check	CG/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP071125

Review By	Rahul	Review On	7/14/2025 11:13:45 AM
Supervise By	Jagrut	Supervise On	7/14/2025 2:30:50 PM
SubDirectory	BP071125	HP Acquire Method	BNA_P
		HP Processing Method	bp070325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025097.D	11 Jul 2025 09:04		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025098.D	11 Jul 2025 10:26		CG/JU	Ok,M
3	PB168787BL	PB168787BL	BP025099.D	11 Jul 2025 11:08		CG/JU	Ok
4	Q2126-02	LOQ-SOIL-02-QT2-202	BP025100.D	11 Jul 2025 11:50	LOQ-SOIL-5 PPM	CG/JU	Ok,M
5	Q2126-02	LOQ-SOIL-02-QT2-202	BP025101.D	11 Jul 2025 12:55	NOT USED	CG/JU	Not Ok
6	Q2126-08	LOQ-WATER-02-QT2-2	BP025102.D	11 Jul 2025 13:36	LOQ-WATER-5 PPM	CG/JU	Ok,M
7	Q2126-08	LOQ-WATER-02-QT2-2	BP025103.D	11 Jul 2025 14:18	LOQ-WATER-10 PPM	CG/JU	Ok
8	Q2126-02	LOQ-SOIL-02-QT2-202	BP025104.D	11 Jul 2025 15:00	LOQ-SOIL-10 PPM	CG/JU	Ok,M
9	Q2126-01	LOD-MDL-SOIL-03-QT	BP025105.D	11 Jul 2025 15:42	LOD-MDL-SOIL-4 PPM	CG/JU	Ok
10	Q2126-01	LOD-MDL-SOIL-03-QT	BP025106.D	11 Jul 2025 16:24	LOD-MDL-SOIL-8 PPM	CG/JU	Ok
11	Q2126-07	LOD-MDL-WATER-01-0	BP025107.D	11 Jul 2025 17:06	LOD-MDL-WATER-4 PPM	CG/JU	Ok
12	Q2126-07	LOD-MDL-WATER-01-0	BP025108.D	11 Jul 2025 17:48	LOD-MDL-WATER-8 PPM	CG/JU	Ok
13	PB168802BL	PB168802BL	BP025109.D	11 Jul 2025 18:30		CG/JU	Ok
14	SSTDCCC040	SSTDCCC040EC	BP025110.D	11 Jul 2025 19:12		CG/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP071525

Review By	Rahul	Review On	7/16/2025 11:48:50 AM
Supervise By	Jagrut	Supervise On	7/16/2025 1:40:00 PM
SubDirectory	BP071525	HP Acquire Method	BNA_P
		HP Processing Method	bp070325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6833,SP6834,SP6835,SP6836,SP6837,SP6838,SP6839,SP6840
CCC	SP6836
Internal Standard/PEM	S12673,10ul/1000ul sample
ICV/I.BLK	SP6770
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP025137.D	15 Jul 2025 08:55		CG/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP025138.D	15 Jul 2025 09:36		CG/JU	Ok,M
3	PB168761TB	PB168761TB	BP025139.D	15 Jul 2025 10:17		CG/JU	Ok
4	PB168787BS	PB168787BS	BP025140.D	15 Jul 2025 10:58		CG/JU	Ok,M
5	Q2517-04MS	TP-14MS	BP025141.D	15 Jul 2025 11:44		CG/JU	Ok
6	Q2517-04MSD	TP-14MSD	BP025142.D	15 Jul 2025 12:25		CG/JU	Ok,M
7	Q2543-01	TP-41	BP025143.D	15 Jul 2025 13:07		CG/JU	Ok
8	Q2565-01	MOO-25-0192-0193	BP025144.D	15 Jul 2025 13:48		CG/JU	Ok,M
9	Q2565-03	MOO-25-0191	BP025145.D	15 Jul 2025 14:30		CG/JU	Ok
10	Q2550-02	VNJ-203	BP025146.D	15 Jul 2025 15:17		CG/JU	Ok,M
11	Q2487-13DL2	G2(0-6)DL2	BP025147.D	15 Jul 2025 15:58		CG/JU	Ok,M
12	Q2597-01	CHRT-26899	BP025148.D	15 Jul 2025 16:39		CG/JU	Ok,M
13	Q2590-01	NB-07-071125	BP025149.D	15 Jul 2025 17:21		CG/JU	Ok,M
14	Q2595-01	CHRT-24849	BP025150.D	15 Jul 2025 18:02		CG/JU	Ok
15	Q2571-01	TP-18	BP025151.D	15 Jul 2025 18:43		CG/JU	Ok
16	Q2539-01	HACKENSACK	BP025152.D	15 Jul 2025 19:25		CG/JU	Ok
17	Q2546-01	HD-02-7-9-2025	BP025153.D	15 Jul 2025 20:06		CG/JU	Ok,M

M : Manual Integration