

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: CAMP02

Lab Code: ACE

SDG NO.: Q2543

Lab File ID: VW031790.D

BFB Injection Date: 07/10/2025

Instrument ID: MSVOA_W

BFB Injection Time: 08:16

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.3
75	30.0 - 60.0% of mass 95	52.2
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.0 (0.0) 1
174	50.0 - 100.0% of mass 95	64.3
175	5.0 - 9.0% of mass 174	4.9 (7.6) 1
176	95.0 - 101.0% of mass 174	63.2 (98.3) 1
177	5.0 - 9.0% of mass 176	3.9 (6.1) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VW031791.D	07/10/2025	08:47
VW0710SBL01	VW0710SBL01	VW031792.D	07/10/2025	10:15
VW0710SBS01	VW0710SBS01	VW031793.D	07/10/2025	11:05
VW0710SBSD01	VW0710SBSD01	VW031794.D	07/10/2025	11:33
TP-41	Q2543-01	VW031809.D	07/10/2025	17:26
TP-49	Q2543-02	VW031810.D	07/10/2025	17:48
TP-23	Q2543-03	VW031811.D	07/10/2025	18:10
TP-23-99	Q2543-04	VW031812.D	07/10/2025	18:32