

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q2552

MATRIX: Water

METHOD: 180.1,SM2130 B

- | | NA | NO | YES |
|--|----|----|-----|
| 1. Blank Contamination - If yes, list compounds and concentrations in each blank: | | ✓ | |
| 2. Sample Duplicate Analysis Met QC Criteria | | | ✓ |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| 3. Digestion Holding Time Met | | ✓ | |
| If not met, list number of days exceeded for each sample: | | | |
| The Holding Times were met for all samples except for WS0725-PT-TURB-WS of Turbidity and for WS0725-PT-TURB-WS of Turbidity as samples were receive out of holding time. | | | |

ADDITIONAL COMMENTS:

The time of sampling was not listed in the COC.

QA REVIEW

Date