

Lab Sample ID:

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

Q2552-02

Report of Analysis

Client: Alliance Technical Group, LLC - Newark Date Collected: 07/07/25 10:00

Project: NJ Drinking Water PT Date Received: 07/09/25

Client Sample ID: WS0725-PT-TURB-WS SDG No.: Q2552

% Solid: 0

Water

Matrix:

Parameter	Conc.	Qua.	DF MD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Turbidity	3.85	Н	1 0.13	1.00	NTU		07/24/25 12:12	SM 2130 B-20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits