



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q2552

Project: NJ Drinking Water PT

RunNo.: LB136603

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Turbidity	ICV	NTU	10.702	10	107	90-110	07/24/2025
Sample ID: Turbidity	CCV1	NTU	9.723	10	97	90-110	07/24/2025
Sample ID: Turbidity	CCV2	NTU	10.150	10	102	90-110	07/24/2025

Initial and Continuing Calibration Verification

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q2552

Project: NJ Drinking Water PT

RunNo.: LB136604

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Turbidity	ICV	NTU	10.702	10	107	90-110	07/24/2025
Sample ID: Turbidity	CCV1	NTU	9.723	10	97	90-110	07/24/2025
Sample ID: Turbidity	CCV2	NTU	10.150	10	102	90-110	07/24/2025