



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Blank Summary

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q2552

Project: NJ Drinking Water PT

RunNo.: LB136603

Analyte		Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: Turbidity	ICB	NTU	0.321	0.5000	J	0.15	1.0	07/24/2025
Sample ID: Turbidity	CCB1	NTU	0.337	0.5000	J	0.15	1	07/24/2025
Sample ID: Turbidity	CCB2	NTU	0.328	0.5000	J	0.15	1	07/24/2025

Initial and Continuing Calibration Blank Summary

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q2552

Project: NJ Drinking Water PT

RunNo.: LB136604

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Turbidity	NTU	0.321	0.5000	J	0.15	1.0	07/24/2025
Sample ID: CCB1 Turbidity	NTU	0.337	0.5000	J	0.15	1	07/24/2025
Sample ID: CCB2 Turbidity	NTU	0.328	0.5000	J	0.15	1	07/24/2025