

Duplicate Sample Summary

Client:	Alliance Technical Group, LLC - Newark	SDG No.:	Q2552
Project:	NJ Drinking Water PT	Sample ID:	Q2552-01
Client ID:	WS0725-PT-TURB-WSDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Turbidity	NTU	+/-20	3.85		3.84		1	0.26		07/24/2025

Duplicate Sample Summary

Client:	Alliance Technical Group, LLC - Newark	SDG No.:	Q2552
Project:	NJ Drinking Water PT	Sample ID:	Q2552-02
Client ID:	WS0725-PT-TURB-WSDUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Turbidity	NTU	+/-20	3.85		3.84		1	0.26		07/24/2025