

metals
- 5a -
MATRIX SPIKE SUMMARY

client: Alliance Technical Group, LLC - Newark **level:** low **sdg no.:** Q2552
contract: ALLI03 **lab code:** ACE
matrix: Water **sample id:** Q2725-02 **client id:** CompMS
Percent Solids for Sample: NA **Spiked ID:** Q2725-02MS **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	2.67		0.26		4.0	60	N	CV

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: Alliance Technical Group, LLC - Newark **level:** low **sdg no.:** Q2552
contract: ALLI03 **lab code:** ACE
matrix: Water **sample id:** Q2725-02 **client id:** CompMSD
Percent Solids for Sample: NA **Spiked ID:** Q2725-02MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Mercury	ug/L	75 - 125	2.32		0.26		4.0	52	N	CV