

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q2552

Contract: ALLI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169068BS Mercury	ug/L	4.0	4.00		100	85 - 115	CV

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q2552

Contract: ALLI03

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB169295BS							
Aluminum	ug/L	2000	1950		98	85 - 115	P
Antimony	ug/L	800	804		100	85 - 115	P
Arsenic	ug/L	800	761		95	85 - 115	P
Barium	ug/L	200	191		96	85 - 115	P
Beryllium	ug/L	200	194		97	85 - 115	P
Boron	ug/L	300	283		94	85 - 115	P
Cadmium	ug/L	200	184		92	85 - 115	P
Calcium	ug/L	1000	1000	J	100	85 - 115	P
Chromium	ug/L	400	393		98	85 - 115	P
Cobalt	ug/L	200	189		94	85 - 115	P
Copper	ug/L	300	301		100	85 - 115	P
Iron	ug/L	3000	2970		99	85 - 115	P
Lead	ug/L	1000	927		93	85 - 115	P
Magnesium	ug/L	2000	1960	J	98	85 - 115	P
Manganese	ug/L	200	199		100	85 - 115	P
Molybdenum	ug/L	400	402		100	85 - 115	P
Nickel	ug/L	500	475		95	85 - 115	P
Potassium	ug/L	10000	9460		95	85 - 115	P
Selenium	ug/L	2000	1940		97	85 - 115	P
Silver	ug/L	75.0	73.2		98	85 - 115	P
Sodium	ug/L	3000	3180		106	85 - 115	P
Thallium	ug/L	2000	1930		96	85 - 115	P
Vanadium	ug/L	300	293		98	85 - 115	P
Zinc	ug/L	200	202		101	85 - 115	P