



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q2552

Contract: ALLI03

Lab Code: ACE

Initial Calibration Source:

Continuing Calibration Source:

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI	Aluminum	19.7	20.0	98	70 - 130	P	08/18/2025	16:21	LB136878
	Antimony	2.10	2.0	105	70 - 130	P	08/18/2025	16:21	LB136878
	Arsenic	1.29	1.0	129	70 - 130	P	08/18/2025	16:21	LB136878
	Barium	9.88	10.0	99	70 - 130	P	08/18/2025	16:21	LB136878
	Beryllium	1.13	1.0	113	70 - 130	P	08/18/2025	16:21	LB136878
	Cadmium	0.81	1.0	81	70 - 130	P	08/18/2025	16:21	LB136878
	Chromium	2.03	2.0	102	70 - 130	P	08/18/2025	16:21	LB136878
	Cobalt	1.07	1.0	107	70 - 130	P	08/18/2025	16:21	LB136878
	Copper	2.04	2.0	102	70 - 130	P	08/18/2025	16:21	LB136878
	Iron	41.2	50.0	82	70 - 130	P	08/18/2025	16:21	LB136878
	Lead	0.84	1.0	84	70 - 130	P	08/18/2025	16:21	LB136878
	Manganese	0.94	1.0	94	70 - 130	P	08/18/2025	16:21	LB136878
	Molybdenum	5.10	5.0	102	70 - 130	P	08/18/2025	16:21	LB136878
	Nickel	0.92	1.0	92	70 - 130	P	08/18/2025	16:21	LB136878
	Selenium	4.91	5.0	98	70 - 130	P	08/18/2025	16:21	LB136878
	Silver	1.08	1.0	108	70 - 130	P	08/18/2025	16:21	LB136878
	Thallium	1.03	1.0	103	70 - 130	P	08/18/2025	16:21	LB136878
	Vanadium	5.06	5.0	101	70 - 130	P	08/18/2025	16:21	LB136878
	Zinc	4.78	5.0	96	70 - 130	P	08/18/2025	16:21	LB136878