



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q2552

Contract: ALLI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	3.80	+/-10	U	20.0	P	08/18/2025	16:02	LB136878
	Antimony	0.32	+/-1	U	2.00	P	08/18/2025	16:02	LB136878
	Arsenic	0.17	+/-0.5	J	1.00	P	08/18/2025	16:02	LB136878
	Barium	0.29	+/-5	U	10.0	P	08/18/2025	16:02	LB136878
	Beryllium	0.26	+/-0.5	U	1.00	P	08/18/2025	16:02	LB136878
	Cadmium	0.14	+/-0.5	U	1.00	P	08/18/2025	16:02	LB136878
	Chromium	0.047	+/-1	U	2.00	P	08/18/2025	16:02	LB136878
	Cobalt	0.050	+/-0.5	U	1.00	P	08/18/2025	16:02	LB136878
	Copper	0.65	+/-1	U	2.00	P	08/18/2025	16:02	LB136878
	Iron	1.96	+/-25	U	50.0	P	08/18/2025	16:02	LB136878
	Lead	0.27	+/-0.5	U	1.00	P	08/18/2025	16:02	LB136878
	Manganese	0.19	+/-0.5	U	1.00	P	08/18/2025	16:02	LB136878
	Nickel	0.15	+/-0.5	U	1.00	P	08/18/2025	16:02	LB136878
	Molybdenum	0.67	+/-2.5	U	5.00	P	08/18/2025	16:02	LB136878
	Selenium	2.86	+/-2.5	U	5.00	P	08/18/2025	16:02	LB136878
	Silver	0.17	+/-0.5	J	1.00	P	08/18/2025	16:02	LB136878
	Thallium	0.11	+/-0.5	U	1.00	P	08/18/2025	16:02	LB136878
	Vanadium	0.11	+/-2.5	U	5.00	P	08/18/2025	16:02	LB136878
	Zinc	0.62	+/-2.5	U	5.00	P	08/18/2025	16:02	LB136878

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Alliance Technical Group, LLC - Newark

SDG No.: Q2552

Contract: ALLI03

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	3.80	+/-10	U	20.0	P	08/18/2025	16:18	LB136878
	Antimony	0.32	+/-1	U	2.00	P	08/18/2025	16:18	LB136878
	Arsenic	0.47	+/-0.5	J	1.00	P	08/18/2025	16:18	LB136878
	Barium	0.29	+/-5	U	10.0	P	08/18/2025	16:18	LB136878
	Beryllium	0.26	+/-0.5	U	1.00	P	08/18/2025	16:18	LB136878
	Cadmium	0.14	+/-0.5	U	1.00	P	08/18/2025	16:18	LB136878
	Chromium	0.047	+/-1	U	2.00	P	08/18/2025	16:18	LB136878
	Cobalt	0.050	+/-0.5	U	1.00	P	08/18/2025	16:18	LB136878
	Copper	0.65	+/-1	U	2.00	P	08/18/2025	16:18	LB136878
	Iron	1.96	+/-25	U	50.0	P	08/18/2025	16:18	LB136878
	Lead	0.27	+/-0.5	U	1.00	P	08/18/2025	16:18	LB136878
	Manganese	0.19	+/-0.5	U	1.00	P	08/18/2025	16:18	LB136878
	Nickel	0.15	+/-0.5	U	1.00	P	08/18/2025	16:18	LB136878
	Selenium	2.86	+/-2.5	U	5.00	P	08/18/2025	16:18	LB136878
	Molybdenum	0.67	+/-2.5	U	5.00	P	08/18/2025	16:18	LB136878
	Silver	0.15	+/-0.5	U	1.00	P	08/18/2025	16:18	LB136878
	Thallium	0.11	+/-0.5	U	1.00	P	08/18/2025	16:18	LB136878
CCB02	Vanadium	0.11	+/-2.5	U	5.00	P	08/18/2025	16:18	LB136878
	Zinc	0.62	+/-2.5	U	5.00	P	08/18/2025	16:18	LB136878
	Aluminum	3.80	+/-10	U	20.0	P	08/18/2025	17:03	LB136878
	Antimony	0.32	+/-1	U	2.00	P	08/18/2025	17:03	LB136878
	Arsenic	0.14	+/-0.5	U	1.00	P	08/18/2025	17:03	LB136878
	Barium	0.29	+/-5	U	10.0	P	08/18/2025	17:03	LB136878
	Beryllium	0.26	+/-0.5	U	1.00	P	08/18/2025	17:03	LB136878
	Cadmium	0.14	+/-0.5	U	1.00	P	08/18/2025	17:03	LB136878
	Chromium	0.047	+/-1	U	2.00	P	08/18/2025	17:03	LB136878
	Cobalt	0.050	+/-0.5	U	1.00	P	08/18/2025	17:03	LB136878
	Copper	0.65	+/-1	U	2.00	P	08/18/2025	17:03	LB136878
	Iron	1.96	+/-25	U	50.0	P	08/18/2025	17:03	LB136878
	Lead	0.27	+/-0.5	U	1.00	P	08/18/2025	17:03	LB136878
	Manganese	0.19	+/-0.5	U	1.00	P	08/18/2025	17:03	LB136878
	Nickel	0.15	+/-0.5	U	1.00	P	08/18/2025	17:03	LB136878
	Molybdenum	0.67	+/-2.5	U	5.00	P	08/18/2025	17:03	LB136878
	Selenium	2.86	+/-2.5	U	5.00	P	08/18/2025	17:03	LB136878
	Silver	0.15	+/-0.5	U	1.00	P	08/18/2025	17:03	LB136878
	Thallium	0.11	+/-0.5	U	1.00	P	08/18/2025	17:03	LB136878
	Vanadium	0.11	+/-2.5	U	5.00	P	08/18/2025	17:03	LB136878
	Zinc	0.62	+/-2.5	U	5.00	P	08/18/2025	17:03	LB136878