

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Alliance Technical Group, LLC -

Contract: ALLI03

Lab Code: ACE

SDG No.: Q2552

Instrument ID: P7

Start Date : 08/18/2025

Run Number: LB136878

End Date : 08/18/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1322	100		100							
S2	S2	1328	99		100							
S3	S3	1331	99		100							
S4	S4	1334	100		102							
S5	S5	1337	103		103							
S6	S6	1340	105		106							
S7	S7	1343	108		109							
S8	S8	1346	111		106							
ICV01	ICV01	1556	104		105							
ICB01	ICB01	1602	103		105							
ICSA01	ICSA01	1605	106		105							
ICSAB01	ICSAB01	1610	110		108							
CCV01	CCV01	1613	114		112							
CCB01	CCB01	1618	104		109							
CRI	CRI	1621	106		107							
PB169290BL	PB169290BL	1624	104		105							
PB169290BS	PB169290BS	1627	112		113							
Q2552-04	WS0725-PT-TM	1641	104		108							
CCV02	CCV02	1649	114		113							
CCB02	CCB02	1703	105		108							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Client: Alliance Technical Group, LLC -

Contract: ALLI03

Lab Code: ACE

SDG No.: Q2552

Instrument ID: P7

Start Date : 08/18/2025

Run Number: LB136878

End Date : 08/18/2025

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1322										
S2	S2	1328										
S3	S3	1331										
S4	S4	1334										
S5	S5	1337										
S6	S6	1340										
S7	S7	1343										
S8	S8	1346										
ICV01	ICV01	1556										
ICB01	ICB01	1602										
ICSA01	ICSA01	1605										
ICSAB01	ICSAB01	1610										
CCV01	CCV01	1613										
CCB01	CCB01	1618										
CRI	CRI	1621										
PB169290BL	PB169290BL	1624										
PB169290BS	PB169290BS	1627										
Q2552-04	WS0725-PT-TM	1641										
CCV02	CCV02	1649										
CCB02	CCB02	1703										