



Lab Name:	<u>Alliance</u>	Contract:	<u>ALLI03</u>	
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2552</u>	
Instrument ID:	<u>ECD_Q</u>	Calibration Date(s):	<u>08/08/2025</u>	<u>08/08/2025</u>
		Calibration Times:	<u>10:48</u>	<u>11:41</u>

LAB FILE ID:	RT 0.1 =	<u>PQ070738.D</u>	RT 0.05 =	<u>PQ070739.D</u>	
RT 0.025 =	<u>PQ070740.D</u>	RT 0.5 =	<u>PQ070741.D</u>	RT 0.25 =	<u>PQ070742.D</u>

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Lab Name:	<u>Alliance</u>	Contract:	<u>ALLI03</u>
Lab Code:	<u>ACE</u>	SDG NO.:	<u>Q2552</u>
Instrument ID:	<u>ECD_Q</u>	Calibration Date(s):	<u>08/08/2025</u> <u>08/08/2025</u>
		Calibration Times:	<u>10:48</u> <u>11:41</u>

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 0.1 = <u>PQ070738.D</u>	RT 0.05 = <u>PQ070739.D</u>
RT 0.025 = <u>PQ070740.D</u>	RT 0.5 = <u>PQ070741.D</u>	RT 0.25 = <u>PQ070742.D</u>

COMPOUND	RT 0.1	RT 0.05	RT 0.025	RT 0.5	RT 0.25	MEAN RT	RT WINDOW FROM TO	
DBCP	4.99	4.99	4.99	4.99	4.99	4.99	4.89	5.09
EDB	1.97	1.98	1.98	1.97	1.97	1.97	1.87	2.07